

Baxter


Rethink Your Hypertension Program

How Blood Pressure Averaging May Improve The Accuracy Of Hypertension Diagnoses

Introduction

Taking a patient's blood pressure (BP) may feel like second nature. It is, after all, one of the most commonly collected vital signs in outpatient and primary care. Although common, inaccurate readings can occur, often resulting from measurement technique or improper patient positioning.¹ Automated oscillometric devices are less prone to human error and can help provide more accurate BP readings.²

But is moving to a digital solution enough? Capturing only one BP reading per exam could be putting your patients at risk for inaccurate hypertension diagnoses. Researchers wanted to better understand the benefits of blood pressure averaging, so they designed a study to compare single in-office blood pressure readings with multiple averaged readings.



Automated devices have started to replace manual auscultation and have even been shown to provide practices with **COST AND TIME-SAVINGS**.³

Methodology

To better understand the benefits of blood pressure averaging, researchers compared the accuracy of single, in-office readings with multiple, averaged readings.

Dr. Robert Smith and his staff at the St. Francis Cardiology Clinic studied the blood pressure readings of 187 adult patients. They used an automated **Welch Allyn** vital signs device to capture patient data. The device's inflation-based **SureBP** algorithm captured multiple blood pressure data points and calculated an average reading for each patient. As part of the algorithm's program, the device only averages readings that are believed to be stable, helping to support accuracy.⁴

Baselining technology – an important component to blood pressure averaging algorithms – was enabled on devices throughout the study. This feature helps account for factors that can cause variability during a blood pressure reading, like cuff placement over clothing, crossed feet, talking during an exam, a distended patient bladder and more.⁴

To recreate the physician office experience, researchers began capturing a patient's first blood pressure measurement at various times after they entered the exam room – ranging from one to three minutes after entry.⁴

Five additional measurements were captured automatically by the device at one-minute intervals after the initial measurement. After collecting multiple blood pressure readings, the device automatically calculated a recommended average that included up to six of the acquired measurements.⁴

Study Parameters⁴



Results

Variability is Common

Across the six readings captured for each patient, researchers found a variability of 15.43 mmHg and a standard deviation of 9.26 mmHg. This means that most patients showed variations greater than 10 mmHg between readings. And in 41% of cases, the first reading was the highest in the series.⁴

One Reading May Be Misleading

If only one blood pressure reading was used as the basis for diagnosis,¹ 27% of patients in this study may have received improper care – 13% of patients would not have gotten the appropriate management and follow-ups, while 14% would have been managed more aggressively than needed.⁴

With a single-reading method:

- 5 patients with an elevated blood pressure – who require more frequent assessment – would have been missed altogether.
- 12 patients who are Stage 1 Hypertensive would have been undiagnosed.
- 27 patients would have been misclassified as Stage 2 Hypertensive.


Help limit reading-to-reading variability with blood pressure averaging for an accurate, holistic picture of your patient's hypertension status.⁴

The American Heart Association recommends taking and averaging at least two blood pressure readings as the preferred method from in office BP measurements¹




3 OUT OF 4 PATIENTS

showed greater than 10 mmHg variability between blood pressure measurements



of patients had a high first-time reading



Conclusion

As automated vital signs solutions have become more common and sophisticated, it's easier than manual measurement to implement blood pressure averaging at your practice. Some solutions— like the **Welch Allyn Spot Vital Signs 4400** device and **Welch Allyn Connex Spot Monitor**— even include blood pressure averaging in the main workflow.

With noticeable differences between individual readings, relying on a single blood pressure measurement can result in misdiagnosis and delay the treatment regimens your patients may need. Measuring and averaging at least two blood pressure readings can help provide an accurate reading, helping you diagnose hypertension with confidence. If your practice doesn't utilize blood pressure averaging, now is the time to start.

Both the **Connex Spot Monitor** and the **Spot Vital Signs 4400** device are simple, easy-to-use solutions that offers an efficient way to capture, access and document vital signs so you can spend more time focused on your patients. It also includes blood pressure averaging right on the home screen for improved hypertension detection and diagnosis.

Begin using blood pressure averaging at your practice today.

Contact your Baxter representative to learn more.

References

1. Writing Committee Members*, Jones DW, Ferdinand KC, Taler SJ, et al. 2025 AHA/ACC/AANP/AAPA/ABC/ACCP/ACPM/AGS/AMA/ASPC/NMA/PCNA/SGIM Guideline for the Prevention, Detection, Evaluation and Management of High Blood Pressure in Adults: A Report of the American College of Cardiology/American Heart Association Joint Committee on Clinical Practice Guidelines. *Circulation*. 2025;152(11):e114-e218. doi:10.1161/CIR.0000000000001356
2. Muntner, Paul et al. "Measurement of Blood Pressure in Humans: A Scientific Statement From the American Heart Association." *Hypertension* 73, no. 5. pg.e35. (May 1, 2019). Doi:10.1161/hyp.0000000000000087.
3. Yarows SA. How to measure blood pressure in primary care offices to assure accuracy while maintaining efficiency. *J Clin Hypertens*. 2017;19(12):1386-1387.
4. Robert Smith, MD. Blood pressure averaging methodology: Decreasing the risk of misdiagnosing hypertension.

Rx Only. For safe and proper use of the product mentioned herein, please refer to the appropriate Operator's Manual or Instructions for Use.

Baxter reserves the right to make changes without notice in design, specifications and models. The only warranty Baxter makes is the express written warranty extended on the sale or rental of its products.

Baxter.com

Baxter, Hillrom, Spot Vital Signs, SureBP, Welch Allyn are trademarks of Baxter International, Inc. or its subsidiaries. US-FLC157-230003 (v2.0) 02/2026

