

Guidance and Manufacturer's Declaration

This guidance and manufacturer's declaration pertains to the Welch Allyn GS 777 Wall Transformer.

EMC Compliance

Special precautions concerning electromagnetic compatibility (EMC) must be taken for all medical electrical equipment. This device complies with IEC 60601-1-2.

- All medical electrical equipment must be installed and put into service in accordance with the EMC information provided in this document and the GS 777 Wall Transformer Directions for Use.
- Portable and mobile RF communications equipment can affect the behavior of medical electrical equipment.

The GS 777 Wall Transformer complies with all applicable and required standards for electromagnetic interference.

- It does not normally affect nearby electronic equipment and devices.
- It is not normally affected by nearby electronic equipment and devices.
- It is not safe to operate the GS 777 Wall Transformer in the presence of high-frequency surgical equipment.
- It is a good practice to avoid using the GS 777 Wall Transformer in extremely close proximity to other electronic equipment.

Emissions and Immunity Information

Electromagnetic Emissions

The GS 777 Wall Transformer is intended for use in the electromagnetic environment specified below. The customer or user of the GS 777 Wall Transformer should assure that it is used in such an environment.

Emissions Test	Compliance	Electromagnetic Environment - Guidance
RF emissions CISPR 11	Group 1	The GS 777 Wall Transformer uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. The GS 777 Wall Transformer is suitable for use in all establishments including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
RF emissions CISPR 11	Class B	
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	


Electromagnetic Immunity

The GS 777 Wall Transformer is intended for use in the electromagnetic environment specified below. The customer or user of the GS 777 Wall Transformer should assure that it is used in such an environment.

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment - Guidance
Electrostatic discharge (ESD) IEC 61000-4-2	± 6 kV contact ± 8 kV air	± 6 kV contact ± 8 kV air	Floors should be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line(s) to line(s) ±2 kV line(s) to earth	±1 kV line(s) to line(s) ±2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions, and voltage variations on power supply input lines. IEC 61000-4-11	>95% dip for 0.5 cycle 60% dip for 5 cycles 30% dip for 25 cycles >95% dip for 5 seconds	>95% dip for 0.5 cycle 60% dip for 5 cycles 30% dip for 25 cycles >95% dip for 5 seconds	Mains power quality should be that of a typical commercial or hospital environment. If the user of the GS 777 Wall Transformer requires continued operation during power mains interruptions, it is recommended that the GS 777 Wall Transformer be powered from an uninterruptible power supply or a battery.
Power frequency (50–60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

Electromagnetic Immunity

The GS 777 Wall Transformer is intended for use in the electromagnetic environment specified below. The customer or user of the GS 777 Wall Transformer should assure that it is used in such an environment.

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment - Guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms 150 kHz to 80 MHz	<p>Portable and mobile RF communications equipment should be used no closer to any part of the GS 777 Wall Transformer than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended separation distance</p> $d = (1.2) \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m 80 MHz to 2.5 GHz	$d = (1.2) \sqrt{P} \text{ 80 MHz to 800 MHz}$ $d = (2.3) \sqrt{P} \text{ 800 MHz to 2.5 GHz}$ <p>where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,^a should be less than the compliance level in each frequency range.^b</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> 

Note 1: At 80 MHz and 800 MHz, the higher frequency range applies.

Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.

^a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the GS 777 Wall Transformer is used exceeds the applicable RF compliance level above, the GS 777 Wall Transformer should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the GS 777 Wall Transformer.

^b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

Recommended Separation Distances Between Portable and Mobile RF Communications Equipment and the GS 777 Wall Transformer

The GS 777 Wall Transformer is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or user of the GS 777 Wall Transformer can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the GS 777 Wall Transformer as recommended below, according to the maximum output power of the communications equipment.

Rated Max. Output Power of Transmitter (W)	Separation Distance According to Frequency of Transmitter (m)		
	150 kHz to 80 MHz $d = (1.2) \sqrt{P}$	80 MHz to 800 MHz $d = (1.2) \sqrt{P}$	800 MHz to 2.5 GHz $d = (2.3) \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

Note 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.

© 2015 Welch Allyn. All rights are reserved. To support the intended use of the product described in this publication, the purchaser of the product is permitted to copy this publication, for internal distribution only, from the media provided by Welch Allyn. No other use, reproduction, or distribution of this publication, or any part of it, is permitted without written permission from Welch Allyn.

DIR 80017722 Ver. B

Welch Allyn, Inc.
4341 State Street Road
Skaneateles Falls, New York 13153-0220
USA

www.welchallyn.com