

January 8, 2014

Dear Valued Customer,

Congratulations on your new Lithium- Ion Phosphate battery replacement for your vital signs device! We are very excited to introduce this new battery technology as it will provide you with improved product performance by increasing battery life, increasing operating capacity of your product and decreasing charging time. Enclosed is the rechargeable, Lithium-Ion Phosphate replacement battery for your Spot (REF 105631) or Spot LXi (REF 105632) Vital Signs device.

The items included in this kit are:

- Lithium-Ion (LFP) Battery
 - Note: In order to install the battery you will need a Phillips screw driver.
- Illustrated battery installation instructions
- Lithium-Ion (Li-ion) battery technology label for you to apply to your vital signs device
- New Directions for use CD (please discard any older versions of the Directions for Use)

Battery Gauge Performance and Charging Behavior:

While charging, the charger icon remains on and the battery icon segments continuously sequence. When the battery is fully charged, all battery icon segments remain on. When the device is not plugged into the wall, it operates on its internal battery.

Note: As the battery charge level decreases the indicator bars turn off left to right similar to the behavior of a cell phone's charging indicators. The vital signs device is usable as long as the battery icon is completely full. If anything less than a full battery icon is displayed, Welch Allyn recommends you plug the device into the wall and recharge the battery until fully charged.

Should you have questions or any difficulties installing your new battery, please call Welch Allyn Customer Service. Visit <http://www.welchallyn.com/regional-service-selector.htm>

Sincerely,



Pooja Sinha
Sr. Marketing Specialist