



[www.mortara.com](http://www.mortara.com)

REF 9515-001-50-ENG Rev F1

# Physician's Guide to

VERITAS WITH ADULT AND PEDIATRIC  
RESTING ECG INTERPRETATION



Copyright © 2015  
by Mortara Instrument, Inc.  
7865 N. 86th Street  
Milwaukee, Wisconsin 53224

*This document contains confidential information that belongs to Mortara Instrument, Inc. No part of this document may be transmitted, reproduced, used, or disclosed outside of the receiving organization without the express written consent of Mortara Instrument, Inc. Mortara is a registered trademark of Mortara Instrument, Inc. VERITAS is a trademark of Mortara Instrument, Inc. V6.0*

**PRECAUTIONS:**

**Some Mortara products are not equipped with the pediatric resting ECG interpretation feature. Refer to user manual for proper instructions and precautions pertaining to equipment use.**

# TABLE OF CONTENTS

---

## RHYTHM STATEMENTS

---

Rhythm Statements and Modifiers (UNIPRO).....	3
Rhythm Statements.....	3
Modifiers .....	4
Modifiers Used with Atrial Fibrillation.....	4
Modifiers Used with Atrial Flutter.....	4
Rhythm Statements and Modifiers (UNIPRO32).....	5
Rhythm Statements.....	5
Modifiers .....	5
Modifiers Used with Atrial Fibrillation.....	5
Modifiers Used with Atrial Flutter.....	5

## ADULT CRITERIA

---

Arm Lead Reversal and Dextrocardia .....	7
Wolff-Parkinson-White .....	8
Atrial Enlargement .....	9
Axis Deviation.....	9
Low Voltage.....	10
S1-S2-S3 Pattern .....	10
Pulmonary Disease .....	11

## ADULT CONDUCTION ABNORMALITIES

---

Right Bundle Conduction .....	13
Left Bundle Conduction .....	14
Non Specific Conduction Abnormality .....	16

## ADULT HYPERTROPHY

---

Right Ventricular Hypertrophy.....	17
Left Ventricular Hypertrophy.....	18

## ADULT MYOCARDIAL INFARCT

---

Myocardial Infarct Discussion.....	21
Anterior Infarct.....	22
Septal Infarct .....	24
Anteroseptal Infarct.....	25
Lateral Infarct .....	26
Anterolateral Infarct .....	27
Inferior Infarct .....	28
Inferior Infarct with Posterior Extension.....	29
Infarct Suppressions .....	29

---

## ADULT ST ELEVATION

---

ST Segment Elevation .....	31
Early Repolarization .....	31
Pericarditis .....	32
Anterior and Septal Epicardial Injury .....	33
Lateral Epicardial Injury .....	35
Inferior Epicardial Injury .....	36

## ADULT ST DEPRESSION

---

ST Depression .....	37
---------------------	----

## ADULT T WAVE ABNORMALITIES

---

T Wave Abnormality, Ischemia .....	39
T Wave Abnormality, Nonspecific .....	41

## ADULT BRUGADA

---

Brugada .....	43
---------------	----

## PEDIATRIC CRITERIA

---

Arm Lead Reversal and Dextrocardia .....	45
Wolff-Parkinson-White .....	45
Atrial Enlargement .....	46
Axis Deviation .....	46

## PEDIATRIC CONDUCTION ABNORMALITIES

---

Right Bundle Conduction .....	47
Left Bundle Conduction .....	47
Ventricular Conduction Delay .....	48

## PEDIATRIC HYPERTROPHY

---

Right Ventricular Hypertrophy .....	49
Left Ventricular Hypertrophy .....	52

PEDIATRIC ST SEGMENT ABNORMALITIES

ST Segment Elevation ..... 55  
 ST Segment Depression ..... 55

PEDIATRIC T WAVE ABNORMALITIES

T Wave Abnormality, Ischemia..... 57

PEDIATRIC TRICUSPID ATRESIA

Tricuspid Atresia ..... 61

PEDIATRIC ENDOCARDIAL CUSHION DEFECT

Endocardial Cushion Defect..... 63

PEDIATRIC ATRIAL SEPTAL DEFECT

Atrial Septal Defect ..... 65

PEDIATRIC QT INTERVAL ABNORMALITIES

QT Prolongation ..... 67  
 QT Shortening ..... 67

PEDIATRIC BRUGADA

Brugada ..... 69

REFERENCE SUMMARY

Age Tables..... 71  
 QRS Axis for Age ..... 71  
 QRS Duration for Age..... 71  
 Prolonged PR Duration, Bradycardia, and Tachycardia for Age ..... 71  
 V6 R/S Amplitude Ratio for Age ..... 71  
 V1/V3R R/S Amplitude Ratio for Age..... 71

VERITAS RESTING ECG INTERPRETATION EVALUATION

Introduction and General Methodology ..... 73

Pediatric Ventricular Hypertrophy ..... 73

Pacemaker Detection ..... 74

Comparison by Categories..... 74

Results ..... 77

Definitions ..... 77

Table 1, Rhythm Criteria Truth Matrices ..... 79

Table 2, Contour Criteria Truth Matrices ..... 81

Table 3, Sensitivity, Specificity and Predictive Accuracies, Rhythm Criteria ..... 84

Table 4, Sensitivity, Specificity and Predictive Accuracies, Contour Criteria ..... 85

Interval Measurements ..... 86

Table 5, Accuracy of Interval Measurements ..... 86

Table 6, Stability of Interval Measurements Against Noise ..... 86

## PREFACE

This guide describes the criteria that the Mortara Instrument VERITAS™ Adult and Pediatric Resting ECG Interpretation algorithm utilizes to analyze and provide interpretive statements for 12-lead ECGs.

Adult criteria are considered for patient ages 16 years and older. Adult descriptions are detailed in the first sections of this guide. Pediatric criteria are considered for patient ages 15 years and younger. Pediatric descriptions are detailed in the last sections of this guide.

Interpretive statements have two components, the actual interpretive text, and the optional reason statement, which immediately follows in each statement in this Physician’s Guide and provides a synopsis of the principle criteria used to reach the specified conclusion. The intention is to provide these reason statements where users find them helpful. They can be omitted on all ECGs via a setup function on the electrocardiograph.

Interpretation of all ECGs proceeds in the sequence of the criteria listing. Ordinarily the last valid statement or conclusion reached within a given section supplants all prior statements.

A condition statement follows each interpretive statement. Conditions and their meanings are listed in the table below:

Condition	Meaning
Normal ECG	Normal
Atypical ECG	An unusual pattern has been observed but has no specific significance.
Borderline ECG	Criteria have limited specificity or prognostic significance or where only minimal criteria are met.
Abnormal ECG	Abnormal
Abnormal Rhythm ECG	Abnormal Rhythm
No Further Interpretation Possible	Upon detecting the phenomenon in question, no further useful interpretation of the record is possible.
No Condition Associated	Used with statement prefixes and suffixes.





# RHYTHM STATEMENTS

---

## Rhythm Statements and Modifiers (UNIPRO)\*

Rhythm statements describe the predominant rhythm in the 10 seconds of analyzed data. A modifier, listed after the rhythm statements, may also be added to more accurately describe the type of rhythm.

### Rhythm Statements

Sinus Tachycardia  
Sinus Rhythm  
Sinus Bradycardia

Atrial Tachycardia (abnormal P axis)  
Atrial Rhythm  
Atrial Bradycardia

Junctional Tachycardia (superior P axis and Short PR)  
Junctional Rhythm  
Junctional Bradycardia

Supraventricular Tachycardia (narrow QRS, regular RR, no P)  
Supraventricular Rhythm  
Supraventricular Bradycardia

Undetermined (regular) rhythm

Atrial fibrillation  
Atrial flutter  
Electronic ventricular pacemaker

\*Electrocardiograph is programmed with the UNIPRO communication protocol.

## Modifiers

- ...with (marked) sinus arrhythmia
- ...with first degree AV block
- ...with short PR interval
- ...with second degree AV block, Mobitz Type (I, II)
- ...with (occasional/frequent) ventricular premature complexes
- ...with (occasional/frequent) ectopic premature complexes
- ...with (occasional/frequent) atrial premature complexes
- ...with (occasional/frequent) supraventricular premature complexes
- ...in a pattern of bigeminy
- ...with marked rhythm irregularity, possible non-conducted PAC, SA block, AV block, or sinus pause.

## Modifiers Used with Atrial Fibrillation

- ...with (rapid/slow) ventricular response
- ...with AV block

## Modifiers Used with Atrial Flutter

- ...with aberrant conduction or ventricular premature complexes
- ...cannot rule out atrial flutter (Regular rate near 150)
- ...electronic (atrial/ventricular) pacemaker
- ...contour analysis based on intrinsic rhythm
- ...intermittent Wolff-Parkinson-White pattern

## Rhythm Statements and Modifiers (UNIPRO32)\*\*

Rhythm statements describe the predominant rhythm in the 10 seconds of analyzed data. A modifier, listed after the rhythm statements, may also be added to more accurately describe the type of rhythm.

### Rhythm Statements

Sinus Tachycardia  
 Sinus Rhythm  
 Sinus Bradycardia

Atrial Tachycardia (abnormal P axis)  
 Atrial Rhythm  
 Atrial Bradycardia

Junctional Tachycardia (superior P axis and Short PR)  
 Junctional Rhythm  
 Junctional Bradycardia

Supraventricular Tachycardia (narrow QRS, regular RR, no P)  
 Supraventricular Rhythm  
 Supraventricular Bradycardia

Undetermined (regular) rhythm

Atrial fibrillation  
 Atrial flutter

Criteria for limits of Tachycardia, Bradycardia as well as PR intervals for age are included in the Pediatric Reference Summary.

### Modifiers

- ...with AV block
- ...with prolonged PR interval for age
- ...with (occasional/frequent) ventricular premature complexes
- ...with (occasional/frequent) ectopic premature complexes
- ...with (occasional/frequent) atrial premature complexes
- ...with (occasional/frequent) supraventricular premature complexes
- ...in a pattern of bigeminy
- ...with marked rhythm irregularity, possible non-conducted PAC, SA block, AV block, or sinus pause.

### Modifiers Used with Atrial Fibrillation

- ...with (rapid/slow) ventricular response
- ...with AV block

### Modifiers Used with Atrial Flutter

- ...with aberrant conduction or ventricular premature complexes
- ...cannot rule out atrial flutter
- ...electronic (atrial/ventricular) pacemaker
- ...contour analysis based on intrinsic rhythm
- ...intermittent Wolff-Parkinson-White pattern

\*\*Electrocardiograph is programmed with the UNIPRO32 communication protocol.



# ADULT CRITERIA

---

## Arm Lead Reversal and Dextrocardia

### Criteria

IF	THEN
No Q in lead I and R amplitude < 150µV in lead I <b>or</b> Q amplitude > 0 in lead I and P axis > 90 and PR duration ≥ 110 ms and QRS axis > 90	PRINT "Arm leads reversed" <b>REASON:</b> <i>Inverted P &amp; QRS in lead I</i>
If above criteria are met and R amplitude < 500 µV in lead V6 and Maximum S amplitude > Maximum R amplitude in lead V6 and P amplitude < 20 µV in lead V6 and P' amplitude < -20 µV in lead V6	PRINT "Dextrocardia" <b>REASON:</b> <i>Inverted P &amp; QRS in V6</i>

### Rationale

Simultaneously negative P and QRS contours in lead 'I' are unlikely in a properly recorded ECG. If, in addition, the QRS has a Qr (or rSr') configuration, the most probable explanation is that the arm leads are reversed or dextrocardia is present. If lead V6 has a typical upright configuration, arm lead reversal is more likely: otherwise, dextrocardia is the remaining plausible explanation.

Although the reason statement for both lead reversal and dextrocardia mentions only the inverted P & QRS, the requirement of Qr/rSr' morphology is important to distinguish these cases from pulmonary disease and right ventricular hypertrophy patterns, where rS configurations are the norm. (Further separation from the latter is ensured by the requirement of an inverted P.)

## Wolff-Parkinson-White

SKIP TEST IF
The test for coupled P wave to QRS is negative
<b>or</b> PR duration > 170 ms
<b>or</b> QRS duration < 100 ms
<b>or</b> Heart rate > 120 BPM
<b>or</b> QRS duration > 200 ms
<b>or</b> PR duration > 100 ms and QRS duration > 160 ms

## Criteria

IF	THEN
PR duration < 140 ms and Delta wave is present in 2 leads	PRINT "Atypical Wolff-Parkinson-White Pattern"
Delta wave is present in 2 leads and R amplitude > S amplitude in V1	PRINT "Type A Wolff-Parkinson-White Pattern"
QRS area ratio $\geq 0.6$ in 2 leads of I/V5/V6 and R duration > 30 ms in V2 <b>or</b> Delta wave is present in 2 leads and PR duration is < 140 ms and R amplitude $\leq$ S amplitude in V1	PRINT "Type B Wolff-Parkinson-White Pattern"

## Atrial Enlargement

### Criteria

IF	THEN
Heart rate < 120 and P amplitude > 250 $\mu$ V in any 1 of leads II/III/aVF/V1/V2	PRINT "Possible right atrial enlargement" <b>REASON:</b> 0.25 mV P wave
Heart rate < 120 and P amplitude > 300 $\mu$ V in any 1 of leads II/III/aVF/V1/V2	PRINT "Right atrial enlargement" <b>REASON:</b> 0.3 mV P wave
P' amplitude < -100 $\mu$ V in V1 or V2 and negative P wave area $\geq$ 400 $\mu$ V/ms in the same lead	PRINT "Possible left atrial enlargement" <b>REASON:</b> -0.1 mV P wave in V1/V2
P' amplitude < -150 $\mu$ V in V1 or V2 and negative P wave area $\geq$ 600 $\mu$ V/ms in the same lead	PRINT "Left atrial enlargement" <b>REASON:</b> -0.15 mV P wave in V1/V2

### Rationale

The criteria are the customary ones. For those records meeting only minimal criteria, the qualifier "possible" is used to convey this information. Right atrial enlargement is not "read" for rates of 120 or above, because it is unclear whether increased P amplitude at elevated rates should be attributed to enlargement.

## Axis Deviation

### Criteria

IF	THEN
QRS axis < -20	PRINT "Moderate Left axis deviation" <b>REASON:</b> QRS axis < -20
QRS axis < -30	PRINT "Abnormal Left axis deviation" <b>REASON:</b> QRS axis < -30
QRS axis > 90	PRINT "Moderate Right axis deviation" <b>REASON:</b> QRS axis > 90
QRS axis > 100	PRINT "Abnormal Right axis deviation" <b>REASON:</b> QRS axis > 100
The total net QRS amplitude in leads I, II, and III is < 33% of the total QRS deflection in leads I, II, and III.	PRINT "Indeterminate axis"

### Rationale

The criteria are more or less conventional. Borderline cases are characterized by the use of the term "moderate." (Axis deviation statements are omitted when subsequently identified diagnostic categories may be regarded as the probable cause of the axis deviation.)

Whenever the net amplitude is a small fraction of the total QRS deflection in each lead, the measurement of axis is lacking in meaning. The term "indeterminate axis" is used to convey this information.

## Low Voltage

SKIP TEST IF
QRS duration $\geq$ 120 ms

### Criteria

IF	THEN
Total QRS deflection < 500 $\mu$ V in all limb leads	PRINT "Low QRS voltage in limb leads" <b>REASON:</b> <i>QRS deflection &lt; 0.5 mV in limb leads</i>
Total QRS deflection < 1000 $\mu$ V in all V leads	PRINT "Low QRS voltage in chest leads" <b>REASON:</b> <i>QRS deflection &lt; 1.0 mV in chest leads</i>
If both of the above are true	PRINT "Low QRS voltage" <b>REASON:</b> <i>QRS deflection &lt; 0.5/1.0 mV in limb/chest leads</i>

## S1-S2-S3 Pattern

### Criteria

IF	THEN
S amplitude > 300 $\mu$ V in I and S amplitude > 400 $\mu$ V in II and S amplitude > 700 $\mu$ V in III <b>or</b> S amplitude > R amplitude in leads I, II & III and S amplitude > 200 $\mu$ V in leads I, II & III and the test for R' is negative in any of these leads and age > 15	PRINT "S1-S2-S3 pattern, consistent with pulmonary disease, RVH, or normal variant"



## Pulmonary Disease

<b>SKIP TEST IF</b>
QRS duration $\geq$ 120 ms

### Criteria

The test for pulmonary disease is based on counting how many of its typical characteristics are present.

One point is awarded for each of

- Right atrial enlargement
- QRS axis  $<$  -30
- QRS axis  $>$  90
- QRS axis is indeterminate
- S1-S2-S3 pattern
- Low voltage in limb leads
- Low voltage in chest leads

Three points are awarded if QRS net amplitude is negative in lead V5 and the R (and R') amplitude in V6  $<$  500  $\mu$ V.

IF	THEN
Cumulative points $>$ 3	PRINT "Consistent with pulmonary disease"

### Rationale

There is room to doubt whether sufficient ECG criteria exist to diagnose pulmonary disease. However, if at least 4 (from a list of 8 distinct) features common to pulmonary disease are present, then the comment "consistent with" seems prudent.



# ADULT CONDUCTION ABNORMALITIES

## Right Bundle Conduction

### Criteria

IF	THEN
R amplitude > 100 $\mu$ V in V1 & V2 and R duration > 20 ms in V1 and V2 and no S in V1 or V2 <b>or</b> R' amplitude > 100 $\mu$ V in V1 & V2 and R' duration > 20 ms in V1 & V2 and no S' in V1 or V2	PRINT "RsR' (QR) in V1/V2 consistent with right ventricular conduction delay"
Either of the above is true and QRS duration > 90 ms and QRS duration < 120 ms and S duration $\geq$ 40 ms in any 2 leads of I/aVL/V4/V5/V6	PRINT "Incomplete right bundle branch block" <b>REASON:</b> 90+ ms QRS duration, terminal R in V1/V2, 40+ ms S in I/aVL/V4/V5/V6
QRS duration $\geq$ 120 ms and S duration $\geq$ 40 ms in any 2 leads of I/aVL/V4/V5/V6 and R duration < 100 ms in any 4 leads of I/aVL/V4/V5/V6 and QRS area > 0 in V1 and V1 does not terminate in S or S' <b>or</b> QRS duration > 105 ms and S duration $\geq$ 60 ms in any 3 leads of I/aVL/V4/V5/V6 and R duration > 60 ms in V1 and QRS area > 0 in V1	PRINT "Right bundle branch block" <b>REASON:</b> 120+ ms QRS duration, upright V1, 40+ ms S in I/aVL/V4/V5/V6
The test for right bundle branch block is positive and R amplitude > 1500 $\mu$ V in V1 and QRS axis > 110	PRINT "Right bundle branch block plus possible Right Ventricular Hypertrophy" <b>REASON:</b> RBBB, 1.5 mV R in V1, RAD

### Rationale

Right bundle branch conduction abnormalities exhibit anterior and rightward directed terminal forces. The rightward force should be noticeably prolonged. Thus, in addition to QRS conducting time criteria, tests are included for a widened terminal R wave in V1 and widened terminal S waves in at least two of the lateral leads. Conventional criteria require QRS widths in excess of 0.12 seconds for bundle branch block. However, very wide lateral S waves, a wide R in an upright V1, and QRS duration > 105 ms will also be read as right bundle branch block by most interpreters. This is the basis of the second portion of the complete right bundle branch block test. Specific criteria for right bundle branch block + right ventricular hypertrophy are also included.

## Left Bundle Conduction

## Criteria

IF	THEN
<p>QRS duration &gt; 105 ms and QRS net amplitude &lt; 0 in V1 &amp; V2 S duration ≥ 80 ms in V1 &amp; V2 and no Q is present in 2 leads of I/V5/V6 and R duration ≥ 60 ms in 2 leads of I/aVL/V5/V6</p>	<p>PRINT "Incomplete left bundle branch block" <b>REASON:</b> 105+ ms QRS duration, 80+ ms Q/S in V1/V2 no Q and 60+ ms R in I/aVL/V5/V6</p>
<p>QRS axis ≤ -45 and R amplitude &gt; Q amplitude in I &amp; aVL and a Q is present in I and S or S' amplitude &gt; R amplitude in II</p>	<p>PRINT "Left anterior fascicular block" <b>REASON:</b> QRS axis ≤ -45, QR in I, RS in II</p>
<p>The test for S1-S2-S3 is negative, and the test for Pulmonary Disease is negative and QRS axis ≥ 110 and R amplitude &gt; Q amplitude in III &amp; aVF and a Q is present in III &amp; aVF</p>	<p>PRINT "Left posterior fascicular block" <b>REASON:</b> QRS axis &gt; 109, inferior Q</p>
<p>The test for Incomplete Left Bundle Branch Block is positive and QRS area ratio &gt; 0.25 in I or V6 and R duration ≥ 100 ms in 1 lead of I/aVL/V6 and QRS duration ≥ 160 ms <b>or</b> QRS duration ≥ 140 ms and the average R duration &gt; 85 ms in I/aVL/V6 <b>or</b> QRS duration ≥ 120 ms and the average R duration &gt; 85 ms in I/aVL/V6 and QRS area ratio &gt; 0.4 in 2 leads of I/aVL/V6</p>	<p>PRINT "Left bundle branch block" <b>REASON:</b> 120+ ms QRS duration, 80+ ms Q/S in V1/V2, 85+ ms R in I/aVL/V6</p>

## Rationale

The meaning of incomplete left bundle branch block beyond describing an ECG pattern is unknown. For this reason the criteria for this statement are narrowly defined, and whenever a specific label such as left anterior fascicular block is available, the term incomplete left bundle branch block is suppressed.

The test for left bundle branch block introduces a measurement called the “QRS area ratio,” which is defined as the ratio of the QRS area (algebraic) to the area of a rectangle defined by QRS onset and offset and the peak positive amplitude. The area ratio is large whenever the QRS is upright and has a wide or notched R wave peak. The thresholds used in the above left bundle branch block tests are empirically determined to correlate with typical left bundle branch block patterns. The area ratio is used in lieu of R duration in order to better discriminate between true left bundle branch block and a monophasic (upright) QRS with nonspecific terminal slurring of the R wave leading to increased QRS duration.

Strict criteria for fascicular blocks are used. This should be noted by readers who use simple axis deviation tests.

## Non Specific Conduction Abnormality

### Criteria

IF	THEN
The test for Right Bundle Branch Block is negative and The test for Incomplete Right Bundle Branch Block is negative and The test for Left Bundle Branch Block is negative and The test for Incomplete Left Bundle Branch Block is negative and The test for left anterior fascicular block is negative and The test for left posterior fascicular block is negative and The test for RSR Pattern is negative and QRS duration > 110 ms	PRINT "Nonspecific intraventricular conduction delay" <b>REASON:</b> <i>110+ ms QRS duration</i>
The test for Right Bundle Branch Block is negative and The test for Left Bundle Branch Block is negative and QRS duration > 130 ms	PRINT "Nonspecific intraventricular conduction block" <b>REASON:</b> <i>130+ ms QRS duration</i>

### Rationale

Intraventricular conduction delay is used to connote QRS widening which does not fit any previously defined pattern, and is not so great as to be considered block.

# ADULT HYPERTROPHY

## Right Ventricular Hypertrophy

SKIP TEST IF
The test for Right Bundle Branch Block is positive
<b>or</b> the test for Left Bundle Branch Block is positive
<b>or</b> age < 16
<b>or</b> S amplitude < 250 $\mu$ V in I
<b>or</b> S amplitude > 1000 $\mu$ V in V1
<b>or</b> QRS axis < 60
<b>or</b> QRS duration > 140 ms and net QRS amplitude < 0 in V1
<b>or</b> Q amplitude > S amplitude and R exists in I

### Criteria

The test for right ventricular hypertrophy is based on counting how many of (or in what degree) its common characteristics are present.

One point is awarded for each of:

- R/R' amplitude > 500  $\mu$ V in V1
- Net QRS amplitude > 0 in V1
- Net QRS amplitude > 500  $\mu$ V in V1
- Net QRS amplitude < 0 and S amplitude > 500  $\mu$ V in V5 or V6
- QRS axis  $\geq$  90
- QRS axis  $\geq$  100
- QRS axis  $\geq$  110
- Possible right atrial enlargement has been called
- S1, S2, S3 is present
- Age > 30
- If Indeterminate Axis is true, no points are given for QRS axis

IF	THEN
Cumulative points > 3	PRINT "Possible right ventricular hypertrophy" <b>REASON:</b> <i>Some/all of: prominent R in V1, late transition, RAD, RAE, SSS</i>
Cumulative points > 5	PRINT "Right ventricular hypertrophy" <b>REASON:</b> <i>Some/all of: prominent R in V1, late transition, RAD, RAE, SSS</i>
The test for possible right ventricular hypertrophy is positive and STJ > STM > STE <b>or</b> one of (STM, STE, and T) < -100 $\mu$ V in V1, V2, and V3 and QRS duration < 120 ms	PRINT "Right ventricular hypertrophy with repolarization abnormality" <b>REASON:</b> <i>Some/all of: prominent R in V1, late transition, RAD, RAE, SSS, right precordial ST depression</i>

**NOTE:** STJ = ST segment amplitude at QRS offset; STM = ST segment amplitude at ST segment midpoint; STE = ST segment amplitude at ST segment endpoint; T = peak of the T wave.

## Left Ventricular Hypertrophy

### Criteria

Tests for left ventricular hypertrophy include various voltage criteria, QRS duration, repolarization abnormalities (strain), and left atrial enlargement (as a correlated factor). To arrive at composite voltage criteria, the common standard criteria are scored by degree of excess over the appropriate threshold. These thresholds depend on the age of the patient, as well as the lead or combination of leads.

SKIP TEST IF
The test for Left Bundle Branch Block is positive <b>or</b> QRS duration > 140 ms and net QRS amplitude < 0 in V1

### Thresholds

AGE	S(V1)	R(V5)	R(Max of V5 or V6) + S(V1)
<20	--	--	--
20-29	3.0 mV	3.0 mV	4.5 mV
30-39	2.4 mV	2.6 mV	4.0 mV
40+	2.4 mV	2.6 mV	3.5 mV

A threshold of 1.1 mV for R (aVL) is used independent of age or sex.



## Voltage Criteria

IF	THEN SCORE
$R/R'(aVL) > 1.1 \text{ mV}$	2 points + 1 point/0.1 mV excess
$S/S'(V1) > \text{threshold}$	2 points + 1 point/0.2 mV excess
$R/R'(V5) > \text{threshold}$	2 points + 1 point/0.2 mV excess
$R/R' (V5/V6)+S/S_(V1) > \text{threshold}$	2 points + 1 point/0.3 mV excess

The measure of QRS conduction time in the context of left ventricular hypertrophy is from QRS onset to the peak negative second derivative after the R peak in V5. Ordinarily, the latter point corresponds to the S nadir:

IF	THEN
Cumulative points > 0	Left ventricular hypertrophy is possible
Cumulative points > 2	Moderate voltage criteria for left ventricular hypertrophy exists
Cumulative points > 4	Voltage criteria for left ventricular hypertrophy are present
Peak 2nd derivative - QRS onset > 68 ms in V5	The conduction time threshold is met
The test for Atrial Fibrillation is negative and (STE < STJ) and (STE < -50 $\mu\text{V}$ ) and (R amplitude > 1100 $\mu\text{V}$ ) in at least 1 lead of I, aVL, V4, V5 and V6 <b>or</b> T amplitude (V1) + T amplitude (V6) > 200 $\mu\text{V}$	Left ventricular hypertrophy exists with repolarization abnormalities
Cumulative points are > 0 and the conduction time threshold is exceeded <b>or</b> the criteria for possible left atrial enlargement are met <b>or</b> left ventricular hypertrophy with repolarization abnormalities exists	Non-voltage criteria for left ventricular hypertrophy are present.
Cumulative points are > 0 and voltage criteria exist for left ventricular hypertrophy	PRINT "Minimal voltage criteria for left ventricular hypertrophy, may be normal variant" <b>REASON:</b> <i>Meets criteria in one of: R(aVL), S(V1), R(V5), R(V5/V6)+S(V1)</i>
Cumulative points are > 2 and voltage criteria exist for left ventricular hypertrophy	PRINT "Moderate voltage criteria for left ventricular hypertrophy, may be normal variant" <b>REASON:</b> <i>Meets criteria in one of: R(aVL), S(V1), R(V5), R(V5/V6)+S(V1)</i>

## Left Ventricular Hypertrophy Criteria (Continued)

IF	THEN
Cumulative points are > 4 and voltage criteria exist for left ventricular hypertrophy	PRINT "Voltage criteria for left ventricular hypertrophy" <b>REASON:</b> <i>Meets criteria in one of: R(aVL), S(V1), R(V5), R(V5/V6)+S(V1)</i>
Non-voltage criteria are met and the test for repolarization abnormalities is negative	PRINT "Possible left ventricular hypertrophy" <b>REASON:</b> <i>Voltage criteria plus LAE or QRS widening</i>
Non-voltage criteria are met and repolarization abnormalities exist	PRINT "Left ventricular hypertrophy with repolarization abnormality" <b>REASON:</b> <i>Voltage criteria plus ST/T abnormality</i>
Cumulative points are > 2 <b>or</b> Non-voltage criteria are met	A flag for Left Ventricular Hypertrophy is set which is used in conjunction with other criteria

### Rationale

ECG criteria for left ventricular hypertrophy are imperfect. The sensitivities of various favorite voltage criteria are not better than 30-40%. Specificities greater than 90% may initially seem sufficient, but application to a general population would evidently generate more false than true positives. The philosophy in the above criteria has been to combine several voltage criteria in order to increase the net sensitivity. In order to minimize the impact of an unavoidable decrease of specificity, records minimally exceeding only one criterion and exhibit no non-voltage criteria are identified as possible normal variants. In all cases, records meeting only voltage criteria are identified as such.

Non-voltage tests for left ventricular hypertrophy include the presence of left atrial enlargement, QRS widening, and repolarization changes. Whenever any of these are present in combination with at least one voltage criterion, a stronger statement is made. A new measure of QRS widening is used in place of intrinsicoid deflection time and/or the total QRS width. Instead, an attempt is made to measure the duration of leftward forces in lead V5. The motivation is to be more sensitive than intrinsicoid timing, while avoiding spurious increases in total QRS duration.

Repolarization changes, for the purpose of identifying non-voltage left ventricular hypertrophy criteria, include depressed, downsloping ST segments in any of the lateral leads, or a T amplitude in V1 greater than that in V6.

# ADULT MYOCARDIAL INFARCT

---

## Myocardial Infarct Discussion

Computer criteria for myocardial infarct depart from standard textbook criteria in greater degree than most electrocardiographers would probably expect. The reason is that conventionally accepted criteria describe stereotypical infarction. When these criteria are applied, they have a high specificity, but a very low sensitivity. For example, a recent review of inferior infarct criteria reported a sensitivity of only 4% using New York Heart Association criteria. In order to achieve more useful results, computer programs must incorporate some of the same unpublished “unconventional” criteria used by experienced ECG interpreters.

Conventional criteria focus on Q wave duration as the primary test for the presence of infarction, and the computer tests that follow naturally retain this focus. The single most important additional criterion, seldom mentioned in reviews in infarction criteria, is a test for repolarization abnormalities characteristic of acute or recent infarction. For example, elevated ST segments and negative T waves are strong indicators of infarction in the presence of otherwise non-diagnostic Q waves. Taking into account these repolarization abnormalities greatly increases both sensitivity and specificity for new or recent infarcts. For old infarcts, the problem is more complex. Gains can be made by considering Q and R wave amplitudes and QRS duration. These factors are quantitatively added by converting to “Q duration equivalents.” Thus for every 30  $\mu\text{V}$  of Q amplitude, 1 ms is added to the actual Q duration to obtain an “equivalent” duration. Likewise, for each 120  $\mu\text{V}$  of R amplitude, 1 ms is subtracted, and for every 4 ms of QRS duration beyond 100 ms, 1 ms is added (up to a maximum correction of 5 ms), or subtracted for durations less than 100 ms. This last factor attempts to exploit the frequent increase in QRS duration concomitant with infarction, whether due to left ventricular hypertrophy, peri-infarction block or other types. To further reduce the impact of a wide, but very small Qs, the equivalent duration is reduced by 1 ms for every  $\mu\text{V}$  that the Q amplitude is short of 100  $\mu\text{V}$ .

Age and sex affect the a priori probability of infarction. These factors are also incorporated by modifying the equivalent Q duration. For males, 1 ms is subtracted from the equivalent Q duration for every two years under the age of 40, up to a maximum correction of 10 ms. Likewise, for females, 1 ms is subtracted for every two years under 50, again up to a maximum of 10 ms.

It should be noted that the above adjustments to the equivalent Q duration are not very large, and should not be expected to cause unreasonable departures from conventional interpretation. Mostly, they can expect to affect the certainty attached to a given interpretation.

With some exception, infarct diagnostic statements are given qualifiers intended to reflect the certainty of the particular interpretation. These qualifiers are:

- Cannot rule out. . . Typical equivalent Q duration 30-34 ms
- Possible. . . Typical equivalent Q duration 35-39 ms
- (Unqualified). . . Typical equivalent Q duration 40+ ms

The presence of repolarization abnormalities characteristic of the infarct can cause the qualifier to be omitted, that is, upgrade to strongest statement.

## Anterior Infarct

<b>Define: Alternate T amplitude =</b>
1. If the test for T' is negative, T - larger of STE or T end
2. If the test for T' is positive, lesser of T & T' - larger of STE or T end

<b>SKIP TEST IF</b>
The test for left bundle branch block is positive <b>or</b> QRS duration > 140 ms and net QRS amplitude < 0 in V1

## Criteria

IF	THEN
STM and STE amplitude > 200 $\mu$ V in V3 and V4 and Alternate T amplitude $\geq$ 0 in V3 and V4	Conditions for a new anterior infarct are present
STM and STE amplitude > 50 $\mu$ V in V3 or V4 and Alternate T amplitude < 0 in V3 or V4	Conditions for a recent anterior infarct are present
Criteria for a new or recent anterior infarct are not met and STM amplitude < 30 $\mu$ V in V3 and V4 and Alternate T amplitude $\geq$ 0 in V3 and V4	Conditions for an old anterior infarct are present
Criteria for a new, recent or old anterior infarct are not met	The age description is "age undetermined"
Equivalent Q duration $\geq$ 30 ms in V2 or V4	Test 1 for anterior infarct is positive
Equivalent Q duration $\geq$ 30 ms in V3 or V5	Test 2 for anterior infarct is positive
Equivalent Q duration $\geq$ 30 ms in V3 and Test 1 for anterior infarct is positive <b>or</b> Equivalent Q duration $\geq$ 30 ms in V4 and Test 2 for anterior infarct is positive <b>or</b> R amplitude < 200 $\mu$ V in V4	PRINT "Cannot rule out anterior infarct" <b>REASON:</b> 30 ms Q wave in V3/V4 or R < 0.2 mV in V4
Equivalent Q duration $\geq$ 35 ms in V3 and Test 1 for anterior infarct is positive and the left ventricular hypertrophy flag is not set <b>or</b> Equivalent Q duration $\geq$ 35 ms in V4 and TEST 2 for anterior infarct is positive	PRINT "Possible anterior infarct" <b>REASON:</b> 35 ms Q wave in V3/V4

## Anterior Infarct Criteria (Continued)

IF	THEN
<p>Equivalent Q duration <math>\geq 40</math> ms in V3 and Test 1 for anterior infarct is positive and the left ventricular hypertrophy flag is not set and the test for low voltage in the chest leads is negative and the test for non-specific intraventricular conduction block is negative</p> <p><b>or</b></p> <p>Equivalent Q duration <math>\geq 40</math> ms in V4 and Test 2 for anterior infarct is positive</p> <p><b>or</b></p> <p>If the test for "Cannot rule out anterior infarct" is positive and either recent or new criteria have been met</p>	<p>PRINT "Anterior infarct"</p> <p><b>REASON:</b> <i>40+ ms Q wave and/or ST/T is abnormality in V3/V4</i></p>

IF	THEN APPEND
Anterior infarct is new	Possibly acute
Anterior infarct is new	Probably recent
The age of the anterior infarct is undetermined	Age undetermined
Anterior infarct is old	Probably old

## Septal Infarct

SKIP TEST IF
<p>The test for left bundle branch block is positive</p> <p><b>or</b></p> <p>the test for anterior infarct is positive and Q amplitude &gt; 0 in V1</p> <p><b>or</b></p> <p>QRS duration &gt; 140 ms and net QRS amplitude &lt; 0 in V1</p>

### Criteria

IF	THEN
STM and STE amplitude > 200 $\mu$ V in V2 and alternate T amplitude $\geq$ 0 in V2	"New" septal infarct is present
STM and STE amplitude > 50 $\mu$ V in V2 and alternate T amplitude < 0 in V2	"Recent" septal infarct is present
Septal infarct is not new or recent and STM amplitude < 50 $\mu$ V in V2 and alternate T amplitude $\geq$ 0 in V2	Septal infarct is "old"
The criteria for a septal infarct have been met and it is neither new, recent, or old	Qualifier "Age undetermined" will be used
Equivalent Q duration $\geq$ 30 ms in V2 <b>or</b> the test for Right Bundle Branch Block is positive and Equivalent Q duration > 20 ms in V2	PRINT "Cannot rule out septal infarct" <b>REASON:</b> 30 ms Q wave in V1/V2
Equivalent Q duration $\geq$ 35 ms in V2 and left ventricular hypertrophy flag is not set	PRINT "Possible septal infarct" <b>REASON:</b> 35 ms Q wave in V1/V2
Equivalent Q duration $\geq$ 40 ms in V2 and the left ventricular hypertrophy flag is not set	PRINT "Septal infarct" <b>REASON:</b> 40+ ms Q wave in V1/V2

IF	THEN APPEND
Septal infarct is new	Possibly acute
Septal infarct is new	Probably recent
The age of the septal infarct is undetermined	Age undetermined
Septal infarct is old	Probably old

## Anteroseptal Infarct

SKIP TEST IF
Positive criteria for a Lateral Infarct exists

## Criteria

IF	THEN
Both an anterior infarct and a septal infarct cannot be ruled out	PRINT "Cannot rule out anteroseptal infarct" <b>REASON:</b> 30 ms Q wave in V1-V4
Anteroseptal infarct cannot be ruled out and if the test for anterior infarct or septal infarct is positive	PRINT "Possible anteroseptal infarct" <b>REASON:</b> 35 ms Q wave in V1-V4
Anteroseptal infarct cannot be ruled out and either an unqualified anterior or septal infarct exists	PRINT "Anteroseptal infarct" <b>REASON:</b> 40+ ms Q wave in V1-V4
A recent septal infarct or anterior infarct has been called	"Probably recent" will be appended to the anteroseptal infarct call
Anterior infarct is not recent and either a new septal infarct or anterior infarct exists	"Possibly acute" is appended to the anteroseptal infarct call
Anterior infarct is not new and the tests for septal infarct age undetermined and/or anterior infarct age undetermined are positive	"Age undetermined" is appended to the anteroseptal infarct call
The tests for both septal infarct old and anterior infarct old are positive	"Probably old" is appended to the anteroseptal infarct call

## Lateral Infarct

## Criteria

IF	THEN
STM AND STE AMP > 200 $\mu$ V in V5 & V6 and STM and STE amplitude > 100 $\mu$ V in I & aVL and Alternate T amplitude $\geq$ 0 in I, aVL, V5 & V6	New lateral infarct is present
STM and STE amplitude > 50 $\mu$ V in I, aVL, V5 or V6 and Alternate T amplitude < 0 in I, aVL, V5 or V6	Recent lateral infarct is called
The criteria for new or recent lateral infarct are not met and STM < 30 $\mu$ V in I, aVL, V5 and V6 and alternate T amplitude > 0 in I, aVL, V5 or V6	Old lateral infarct is present
The tests for new, recent or old lateral infarct are negative	Qualifier "age undetermined:" will be used
Equivalent Q duration $\geq$ 30 ms in 2 leads of I/V5/V6 and a lateral infarct cannot be ruled out	PRINT "Cannot rule out lateral infarct" <b>REASON:</b> 30 ms Q wave in I/aVL/V5/V6
Equivalent Q duration $\geq$ 35 ms in 1 lead of I/V5/V6 and a lateral infarct "cannot be ruled out"	PRINT "Possible Lateral infarct" <b>REASON:</b> 35 ms Q wave in I/V5/V6
Equivalent Q duration $\geq$ 40 ms in 1 lead of I/V5/V6 and the test for lateral infarct is positive <b>or</b> lateral infarct "cannot be ruled out" and the tests for a new or recent lateral infarct are positive	PRINT "Lateral infarct" <b>REASON:</b> 40+ ms Q wave and/or ST/T abnormality in I/aVL/V5/V6

IF	THEN APPEND
Lateral infarct is new	Possibly acute
Lateral infarct is recent	Probably recent
The age of the lateral infarct is undetermined	Age undetermined
Lateral infarct is old	Probably old



## Anterolateral Infarct

### Criteria

IF	THEN
Both an anterior infarct and a lateral infarct "cannot be ruled out"	PRINT "Cannot rule out anterolateral infarct" <b>REASON:</b> 30 ms Q wave in I/aVL/V3-V6
The tests for anterior infarct or lateral infarct are positive	PRINT "Possible anterolateral infarct" <b>REASON:</b> 35 ms Q wave in I/aVL/V3-V6
The tests for an unqualified anterior infarct or lateral infarct are positive	PRINT "Anterolateral infarct" <b>REASON:</b> 40+ ms Q wave in I/aVL/V3-V6
The test for either a recent lateral infarct or anterior infarct is positive	"Probably recent" is appended to the anterolateral infarct statement
The infarct is not a recent anterolateral infarct and the test for a new lateral infarct or anterior infarct is positive	"New" anterolateral infarct is present "Probably acute" is appended to the statement
The infarct is not a new anterolateral infarct and the tests for "age undetermined" lateral infarct and/or an "age undetermined" anterior infarct are positive	"Age undetermined" anterolateral infarct is present "Age undetermined" is appended to the record
Both the lateral infarct and anterior infarct are qualified as "old"	Anterolateral infarct call will be qualified as "probably old"

## Inferior Infarct

## Criteria

IF	THEN
STM and STE amplitude > 100 $\mu$ V in II and aVF and Alternate T amplitude $\geq$ 0 in II & aVF	"New" inferior infarct is present
STM and STE amplitude > 50 $\mu$ V in II or aVF and Alternate T amplitude < 0 in II or aVF	"Recent" inferior infarct is present
Inferior infarct is not new or recent and STM amplitude < 30 $\mu$ V in II and aVF and Alternate T amplitude $\geq$ 0 in II and aVF	Inferior infarct is "old"
inferior infarct is not new, recent, or old	Qualifier "age undetermined" will be used
Equivalent Q duration $\geq$ 30 ms in II or aVF Q amplitude in lead I < Q amplitude in lead II <b>or</b> Q amplitude in lead I < Q amplitude in aVF	PRINT "Cannot rule out inferior infarct" <b>REASON:</b> 30 ms Q wave in II/aVF
Equivalent Q duration $\geq$ 35 ms in II or aVF and an inferior infarct cannot be ruled out	PRINT "Possible inferior infarct" <b>REASON:</b> 35 ms Q wave in II/aVF
Inferior infarct cannot be ruled out and Equivalent Q duration $\geq$ 40 ms in II or aVF <b>or</b> The test for a new or recent Inferior infarct is positive	PRINT "Inferior infarct" <b>REASON:</b> 40+ ms Q wave and/or ST/T abnormality in II/aVF
QA > S amplitude in 1 lead of II & aVF	Suppress Abnormal left Axis Deviation
Inferior infarct is "new"	Append "possibly acute"
Inferior infarct is "recent"	Append "probably recent"
Age of the inferior infarct is undetermined	Append "age undetermined"
Inferior infarct is "old"	Append "probably old"

## Inferior Infarct with Posterior Extension

SKIP TEST IF
The test for an inferior infarct is negative
The test for Right Bundle Branch Block is positive
A Q-wave is present in V1 or V2

### Criteria

IF	THEN
R duration $\geq$ 40 ms in V1 & V2 <b>or</b> R duration $\geq$ 35 ms and QRS net amplitude $>$ 0 in V1 or V2 <b>or</b> R duration $\geq$ 30 ms and QRS net amplitude $>$ 0 in V1 and V2	Append "with posterior extension" "prominent R Wave in V1/V2" to the inferior infarct statement

## InfarctSuppressions

### Criteria

IF	THEN
The test for inferior infarct, lateral infarct, anteroseptal infarct or septal infarct is positive	Suppress left axis deviation, incomplete left bundle branch block, intraventricular conduction delay
The test for anteroseptal infarct or a lateral infarct is positive	Suppress pulmonary disease



# ADULT ST ELEVATION

## ST Segment Elevation

SKIP TEST IF
The test for either right bundle branch block, left bundle branch block, intraventricular conduction block, myocardial infarct or left ventricular hypertrophy with repolarization is positive

### Criteria

IF	THEN
STJ/STM/STE all $\geq 50 \mu\text{V}$ and T is not upward inflected in 2 leads of I, II, III, aVF, V3-V6	PRINT "Nonspecific ST elevation" <b>REASON:</b> <i>0.05+ mV ST elevation</i>

## Early Repolarization

SKIP TEST IF
Corrected QT interval $> 450 \text{ ms}$
Either myocardial infarct, right bundle branch block, left bundle branch block, intraventricular conduction block is present and the left ventricular hypertrophy flag is not set

### Criteria

IF	THEN
Count of leads V1-V6 for which STJ and STM amplitude $> 75 \mu\text{V}$ plus count of leads I, II, III, aVL, aVF for which STJ & STM $> 50 \mu\text{V}$ exceeds 2 and sum of STJ amplitudes $> 450 \mu\text{V}$ for leads passing above test	PRINT "ST elevation, consistent with epicardial injury, pericarditis, or early repolarization" <b>REASON:</b> <i>ST elevation w/o normally leads inflected T wave</i>
ST elevation is present, per the above conditions and more than 1/2 of the leads passing ST elevation test above also have well-inflected T waves	PRINT "ST elevation, probably early repolarization" <b>REASON:</b> <i>ST elevation with normally inflected T wave</i>
Above count $> 5$ and sum $> 450 \mu\text{V}$	PRINT "Early repolarization" <b>REASON:</b> <i>ST elevation with normally inflected T wave</i>

## Pericarditis

**SKIP TEST IF**

The test for myocardial infarct, right bundle branch block, left bundle branch block, intraventricular conduction block, left ventricular hypertrophy is positive

**Criteria**

<b>IF</b>	<b>THEN</b>
4 times STJ & T amplitude & T amplitude > 0 in at least 4 leads of I, II, V4-V6 and STJ and STM amplitude > -100 $\mu\text{V}$ in all leads except aVR and count of leads I, II, aVF with STJ and STM amplitude > 75 $\mu\text{V}$ plus count of leads V2-V6 with STJ and STM amplitude > 90 $\mu\text{V}$ is $\geq$ to 5	PRINT "Possible acute pericarditis" <b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave</i>
Possible acute pericarditis is present and count of leads I, II, aVF with STJ and STM amplitude > 90 $\mu\text{V}$ plus count of leads V2-V6 with STJ and STM amplitude > 110 $\mu\text{V}$ is $\geq$ to 5	PRINT "Acute pericarditis" <b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave</i>

## Anterior and Septal Epicardial Injury

### SKIP TEST IF

The test for pericarditis, left bundle branch block, right bundle branch block, intraventricular conduction block is positive

### DEFINE

ST LIMIT = 300  $\mu$ V

(add 100  $\mu$ V for any precordial lead with net QRS amplitude < 0)

### Criteria

IF	THEN
<p>6 times STJ amplitude &gt; QRS deflection in V1 &amp; V2</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT/2</p> <p>and T is not upward inflected in V1 and V2</p> <p>and Alternate T amplitude <math>\geq</math> in V1 and V2</p> <p>and the test for septal infarct is negative</p> <p>and the left ventricular hypertrophy flag is not set</p>	<p>PRINT "Possible septal epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V1/V2</i></p>
<p>4 times STJ amplitude &gt; QRS deflection in V1 &amp; V2</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT</p> <p>and T is not upward inflected in V1 and V2</p> <p>and Alternate T amplitude <math>\geq</math> 0 in V1 and V2</p> <p>and test for septal infarct is negative</p> <p>and the left ventricular hypertrophy flag is not set</p>	<p>PRINT "Septal epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V1/V2</i></p>

## Anterior and Septal Epicardial Injury Criteria (Continued)

IF	THEN
<p>6 times STJ amplitude &gt; QRS deflection in 2 leads of V2-V5</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT/2</p> <p>and T is not upward inflected in 2 leads of V2-V5</p> <p>and Alternate T amplitude <math>\geq 0</math> in V2-V5</p> <p>and the test for anterior infarct is negative</p> <p>and the left ventricular hypertrophy flag is not set</p>	<p>PRINT "Possible anterior epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V2-V5</i></p>
<p>4 times STJ amplitude &gt; QRS deflection in 2 leads of V2-V5</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT</p> <p>and T is not upward inflected in 2 leads of V2-V5</p> <p>and alternate T amplitude <math>\geq 0</math> in V2-V5</p> <p>and the test for anterior infarct is negative</p> <p>and the left ventricular hypertrophy flag is not set</p>	<p>PRINT "Anterior epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V2-V5</i></p>
<p>The test for a possible anterolateral epicardial injury is positive</p> <p>and the test for possible anterior and possible septal epicardial injury is positive</p>	<p>PRINT "Possible anteroseptal epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V1-V4</i></p>
<p>The test for a possible anteroseptal epicardial injury is positive</p> <p>and additional criteria substantiates an anterior injury or septal injury</p>	<p>PRINT "Anteroseptal epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V1-V4</i></p>



## Lateral Epicardial Injury

SKIP TEST IF
The test for pericarditis, left bundle branch block, right bundle branch block, intraventricular conduction block is positive

## Criteria

IF	THEN
<p>6 times STJ amplitude &gt; QRS deflection in 4 leads of I/aVL/V5/V6</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT/2</p> <p>and T is not upward inflected in 2 leads of I, aVL, V5, and V6</p> <p>and Alternate T amplitude <math>\geq 0</math> in I, aVL, V5 and V6</p> <p>and the test for a lateral infarct is negative</p>	<p>PRINT "Possible lateral epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in I/aVL/V5/V6</i></p>
<p>4 times STJ amplitude &gt; QRS deflection in 2 leads of I/aVL/V5/V6</p> <p><b>or</b></p> <p>STJ amplitude &gt; ST LIMIT</p> <p>and T is not upward inflected in 2 leads of I, aVL, V5 and V6</p> <p>and Alternate T amplitude <math>\geq 0</math> in I, aVL, V5 and V6</p> <p>and the test for a lateral infarct is negative</p>	<p>PRINT "Lateral epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in I/aVL/V5/V6</i></p>
<p>The test for both possible anterior and lateral injury is positive</p>	<p>PRINT "Possible anterolateral epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V3-V6</i></p>
<p>The test for possible anterolateral epicardial injury is positive</p> <p>and the test for anterior and/or lateral injury is positive</p>	<p>PRINT "Anterolateral epicardial injury"</p> <p><b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in V3-V6</i></p>

## Inferior Epicardial Injury

### SKIP TEST IF

The test for pericarditis, left bundle branch block, right bundle branch block, intraventricular conduction block is positive

### Criteria

IF	THEN
6 times STJ amplitude > QRS deflection in II and aVF <b>or</b> STJ amplitude > ST LIMIT/2 and T is not upward inflected in II and aVF and Alternate T amplitude $\geq 0$ in II and aVF and the test for inferior infarct is negative	PRINT "Possible inferior epicardial injury" <b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in II/aVF</i>
4 times STJ amplitude > deflection in II & aVF <b>or</b> STJ amplitude > ST LIMIT and T is not upward inflected in II and aVF and Alternate T amplitude $\geq 0$ in II and aVF and the test for inferior infarct is negative	PRINT "Inferior epicardial injury" <b>REASON:</b> <i>Marked ST elevation w/o normally inflected T wave in II/aVF</i>

# ADULT ST DEPRESSION

## ST Depression

SKIP TEST IF
The test for left bundle branch block, intraventricular conduction block, left ventricular hypertrophy with repolarization, or pericarditis is positive

### Criteria

IF	THEN
The tests for right ventricular hypertrophy with repolarization, ST elevation and right bundle branch block are negative and STJ amplitude < -100 $\mu$ V and STE amplitude $\geq 0$ in 2 Leads (except aVR and III)	PRINT "Junctional depression, probably normal" <b>REASON:</b> <i>0.1+ mV junctional ST depression</i>
The tests for right ventricular hypertrophy with repolarization, ST elevation, and right bundle branch block are negative and STJ < -100 $\mu$ V and STE < 0 and STE $\geq$ STJ/2 in 2 leads (except aVR and III)	PRINT "Abnormal Junctional depression" <b>REASON:</b> <i>Junctional depression with weak upslope</i>
The tests for right ventricular hypertrophy with repolarization, ST elevation, and right bundle branch block are negative and STM or STE < both STJ and -50 $\mu$ V in 2 leads (except aVR and III)	PRINT "ST depression, possible digitalis effect" <b>REASON:</b> <i>Downsloping or coved ST depression</i>
The tests for right ventricular hypertrophy with repolarization, ST elevation, and right bundle branch block are negative and STJ/STM/STE all < -25 $\mu$ V in 2 leads (except aVR and III)	PRINT "Minimal ST depression" <b>REASON:</b> <i>0.025+ mV ST depression</i>

## ST Depression Criteria (Continued)

IF	THEN
<p>The tests for right ventricular hypertrophy with repolarization, ST elevation and right bundle branch block are negative and STM &lt; -50 <math>\mu</math>V and STE &lt; 0 or STE &lt; all of STJ/STM -50 <math>\mu</math>V in 2 leads (except aVR and III)</p>	<p>PRINT "Moderate ST depression" <b>REASON:</b> 0.05+ mV ST depression</p>
<p>STJ/STM/STE all &lt; -100 <math>\mu</math>V in 2 leads (except aVR and III) and except V1/V2 if right bundle branch block is present or right ventricular hypertrophy with repolarization is present)</p>	<p>PRINT "Marked ST depression, possible subendocardial injury" <b>REASON:</b> 0.1+ mV ST depression</p>
<p>STJ/STM/STE all &lt; -200 <math>\mu</math>V in 2 leads (except aVR and III and except V1/V2 if right bundle branch block is present or right ventricular hypertrophy with repolarization is present)</p>	<p>PRINT "Marked ST depression, consistent with subendocardial injury" <b>REASON:</b> 0.2+ mV ST depression</p>
<p>The test for atrial fibrillation is positive and either minimal or moderate ST depression is present or "marked" ST depression w/o .1 + mV ST depression</p>	<p>Append "probably digitalis effect"</p>
<p>The test for atrial fibrillation is positive and there is "marked" ST depression w/o .2 + mV ST depression</p>	<p>Append "or digitalis effect"</p>

# ADULT T WAVE ABNORMALITIES

## T Wave Abnormality, Ischemia

SKIP TEST IF
Left bundle branch block, intraventricular conduction block, left ventricular hypertrophy with repolarization, right ventricular hypertrophy with repolarization, subendocardial injury, ST elevation or pericarditis is (are) true

### Criteria

IF	THEN
The test for anteroseptal infarct is negative and the test for right ventricular hypertrophy with repolarization is negative and Alternate T amplitude $\leq -100 \mu\text{V}$ in 2 leads of V2/V3/V4 (excluding V2 if right bundle branch block is present)	PRINT "T wave abnormality, possible anterior ischemia" <b>REASON:</b> <i>-0.1+ mV T wave in V3/V4</i>
The test for anterior ischemia is positive and Alternate T amplitude $< -500 \mu\text{V}$ in 1 lead of V2/V3/V4 (excluding V2 if right bundle branch block is present)	PRINT "T wave abnormality, consistent with anterior ischemia" <b>REASON:</b> <i>-0.5+ mV T wave in V3/V4</i>
The test for lateral infarct is negative and Alternate T amplitude $< -100 \mu\text{V}$ in 2 leads of I/aVL/V4/V5/V6 (excluding aVL if $R(aVL) \leq 500 \mu\text{V}$ )	PRINT "T wave abnormality, possible lateral ischemia" <b>REASON:</b> <i>-0.1+ mV T wave in I/aVL/V5/V6</i>
The test for lateral ischemia is positive and Alternate T amplitude $\leq -500 \mu\text{V}$ in 1 lead of I/aVL/V5/V6 (excluding aVL if $R(aVL) \leq 500 \mu\text{V}$ )	PRINT "T wave abnormality, consistent with lateral ischemia" <b>REASON:</b> <i>-0.5+ mV T wave in I/aVL/V5/V6</i>
The tests for both possible anterior and lateral ischemia are positive	PRINT "T wave abnormality, possible anterolateral ischemia" <b>REASON:</b> <i>-0.1+ mV T wave in V3-V6</i>
The test for possible anterolateral ischemia is positive and lateral and/or anterior ischemia is marked	PRINT "T wave abnormality, consistent with anterolateral ischemia" <b>REASON:</b> <i>-0.5+ mV T wave in I/aVL/V3-V6</i>

## T Wave Abnormality Ischemia Criteria (Continued)

IF	THEN
The test for nonspecific ST abnormalities is positive and the test for possible anterior ischemia and/or possible lateral ischemia is positive	Prefix "ST &" to the T wave abnormality statement
The test for atrial fibrillation is positive and the tests for possible anterior ischemia and/or possible lateral ischemia are positive	Append "or digitalis effect"
The test for inferior infarct is negative and alternate T amplitude < -100 $\mu$ V in II or aVF (excluding aVF if net QRS amplitude < 0) and alternate T amplitude < 0 in II and aVF	PRINT "T wave abnormality, possible inferior ischemia" <b>REASON:</b> <i>-0.1+ mV T wave in II/aVF</i>
The test for inferior ischemia is positive and non-specific ST abnormalities are present	Prefix "ST &" T wave abnormality, possible inferior ischemia
T wave abnormality is present and the test for possible inferior ischemia is positive and the test for possible atrial fibrillation is positive	Append "or digitalis effect" to the T wave abnormality statement
The test for possible inferior ischemia is positive and Alternate T amplitude < -500 $\mu$ V in II or aVF (excluding aVF if net QRS amplitude < 0)	PRINT "T wave abnormality, consistent with inferior ischemia" <b>REASON:</b> <i>-0.5+ mV T wave in II/aVF</i>

## T Wave Abnormality, Nonspecific

**SKIP TEST** (except test Short QT) **IF**

left bundle branch block, intraventricular conduction block, left ventricular hypertrophy with repolarization, right ventricular hypertrophy with repolarization, subendocardial injury, ST elevation, pericarditis, myocardial infarct, right bundle branch block, possible anterior ischemia, possible lateral ischemia or possible inferior ischemia exist

**DEFINE**

TMIN=

1. 25  $\mu$ V = net QRS amplitude/20 if net amplitude > 0
2. 25  $\mu$ V if net amplitude < 0

**Criteria**

IF	THEN
QRS axis - T axis > 60 and T axis < 0 <b>or</b> QRS - T axis < -60 and T axis > 90	PRINT "Abnormal QRS-T angle" <b>REASON:</b> <i>QRS-T axis difference &gt; 60</i>
Count of I/II/aVL/aVF/V3-V6 with alternate T amplitude < TMIN and R amplitude > 500 $\mu$ V is $\geq$ 2	PRINT "Nonspecific T wave abnormality"
Nonspecific ST abnormalities and nonspecific T-wave abnormalities exist and the test for tall T waves is negative	PRINT "Nonspecific ST & T wave abnormality"
The test for atrial fibrillation is positive and the test for either nonspecific T wave or ST abnormalities is positive	Append "probably digitalis effect"
T amplitude > 1000 $\mu$ V and T amplitude > 1/2 R amplitude in 3 leads of I/II/V1-V6	PRINT "Tall T waves, possible hyperkalemia"
QTc < 360 ms and heart rate < 140	PRINT "Short QT interval"
QTc > 450 ms	PRINT "Long QT interval"





# ADULT BRUGADA

---

## Brugada

SKIP TEST IF
QRS duration $\geq$ 110 ms

### Criteria

IF	THEN
STJ $\geq$ 100 $\mu$ V and the lesser of STM or STE $<$ STJ and T amplitude $>$ the lesser of STM or STE in some of V1/V2/V3 and T amplitude $<$ 0 in some of V1/V2/V3	PRINT "Type 3 Brugada pattern (non-diagnostic)"
STJ $\geq$ 200 $\mu$ V and the lesser of STM or STE $<$ STJ and the lesser of STM or STE $>$ 100 $\mu$ V and T amplitude $>$ the lesser of STM or STE in some of V1/V2/V3	PRINT "Type 2 Brugada pattern (non-diagnostic)"
STJ $\geq$ 200 $\mu$ V and the lesser of STM or STE $<$ STJ and T amplitude $<$ 0 in some of V1/V2/V3	PRINT "Type 1 Brugada pattern (non-diagnostic)"



## PEDIATRIC CRITERIA

---

### Arm Lead Reversal and Dextrocardia

#### Criteria

IF	THEN
No Q in lead I and R amplitude < 150 $\mu$ V in lead I  <b>or</b>  Q amplitude > R amplitude in lead I and Maximum S amplitude > 150 $\mu$ V in lead I and P amplitude in lead III > P amplitude lead II and P axis > 90 and QRS axis > 90	PRINT "Arm leads reversed"  <b>REASON:</b> <i>rS or Qr in I, P(III) &gt; P(II), QRS axis &gt; 90</i>

IF	THEN
If P axis $\geq$ 90 and P axis $\leq$ 180 and Maximum R amplitude < 500 $\mu$ V in V6 and Maximum S amplitude > R amplitude in V6 and P amplitude < 20 $\mu$ V in lead V6 and P' amplitude < -20 $\mu$ V in lead V6	PRINT "Dextrocardia"

### Wolff-Parkinson-White

#### Criteria

IF	THEN
If Delta wave is present in some of V1/V2/V3/V4/V5/V6 and PR duration $\leq$ 119 ms and QRS duration $\geq$ 97 ms and Ventricular rate < 150 bpm	PRINT "Wolff-Parkinson-White Syndrome"  <b>REASON:</b> <i>Delta Waves</i>

## Atrial Enlargement

### Criteria

IF	THEN
Age $\geq$ 10 years and P amplitude $>$ 200 $\mu$ V in any 1 lead of I/II/III/aVF/V1/V2 and P amplitude $>$ 150 $\mu$ V in any 2 leads of I/II/III/aVF/V1/V2	PRINT "Consider right atrial enlargement" <b>REASON:</b> 0.2 mV P wave, Age $\geq$ 10 yr
P amplitude $>$ 250 $\mu$ V in any 1 lead of I/II/III/aVF/V1/V2 and P amplitude $>$ 200 $\mu$ V in any 2 leads of I/II/III/aVF/V1/V2	PRINT "Right atrial enlargement" <b>REASON:</b> 0.25 mV P wave
P' amplitude $<$ -70 $\mu$ V and negative P wave area $\geq$ 400 $\mu$ V ms in V1	PRINT "Consider left atrial enlargement"
P' amplitude $<$ -100 $\mu$ V and negative P wave area $\geq$ 400 $\mu$ V/ms in V1	PRINT "Left atrial enlargement"

## Axis Deviation

### Criteria

IF	THEN
QRS axis $<$ Minimum QRS axis for age	PRINT "Left axis deviation for age" <b>REASON:</b> QRS axis $<$ [Minimum QRS axis for age]
QRS axis $>$ Maximum QRS axis for age	PRINT "Right axis deviation" <b>REASON:</b> QRS axis $>$ Maximum QRS axis for age

Please see pediatric criteria table for **QRS Axis for Age** in Pediatric Reference Summary. Axis deviation statements are omitted when subsequently identified diagnostic categories may be regarded as the probable cause of the axis deviation, e.g. right or left bundle branch conduction blocks.

# PEDIATRIC CONDUCTION ABNORMALITIES

## Right Bundle Conduction

### Criteria

IF	THEN
QRS duration $\geq$ Maximum QRS duration for age and R' amplitude $\geq$ 150 $\mu$ V in V1 and R' duration $\geq$ 20 ms in V1 and R' amplitude $>$ 4 x S' amplitude in V1	PRINT "Right bundle branch block" <b>REASON:</b> QRS $\geq$ [Maximum QRS duration for age], RSR' in V1
QRS duration $\geq$ Maximum QRS duration for age and R amplitude $\geq$ 550 $\mu$ V and no S wave is present in V1	PRINT "Right bundle branch block" <b>REASON:</b> QRS $\geq$ [Maximum QRS duration for age], no S in V1

Please see pediatric criteria table for **QRS Duration for Age** in Pediatric Reference Summary.

## Left Bundle Conduction

### Criteria

IF	THEN
QRS axis $\leq$ -60	PRINT "Left anterior superior fascicular block" <b>REASON:</b> QRS axis -60 to -90
S duration $\leq$ 20 ms in 3 of I/aVL/V5/V6 and Terminal QRS axis $\leq$ 90 and QRS duration $\geq$ Maximum QRS duration for age and R wave amplitude $\leq$ 450 $\mu$ V and R wave duration $\leq$ 39 ms in some of V1/V2/V3  <b>or</b>  R wave amplitude $\geq$ 450 $\mu$ V and R wave duration $\leq$ 39 ms in some of V1/V2/V3 and QRS duration $>$ 135 ms	PRINT "Left bundle branch block" <b>REASON:</b> QRS $\geq$ [Maximum QRS duration for age], terminal QRS leftward

Please see pediatric criteria table for **QRS Duration for Age** in Pediatric Reference Summary.

## Ventricular Conduction Delay

### Criteria

IF	THEN
The test for Right Bundle Branch Block is negative and The test for Left Bundle Branch Block is negative and The test for Left Anterior Fascicular Block is negative and QRS duration $\geq$ Maximum QRS Duration for age	PRINT "Ventricular Conduction Delay"  <b>REASON:</b> <i>QRS duration <math>\geq</math> [Maximum QRS Duration for age]</i>

Please see pediatric criteria table for **QRS Duration for Age** in Pediatric Reference Summary.

# PEDIATRIC HYPERTROPHY

## Right Ventricular Hypertrophy

SKIP TEST IF
Test for Right Bundle Branch Block is positive, or Test for Left Bundle Branch Block is positive, or Test for Ventricular Conduction Delay is positive

### Criteria

Criteria statements for Right Ventricular Hypertrophy are printed only if the “Print reason” option on the electrocardiograph is turned on; otherwise, only the summary statements are printed.

IF	THEN
Age $\geq$ 1 month and S amplitude $\geq$ 1000 $\mu$ V in V6	PRINT REASON “RVH voltage criteria: S(V6) > 1mV, 1mo-15yr” <b>NOTE:</b> Final comment is “Borderline ECG”
Age $\geq$ 1 month and R amplitude $\geq$ 2500 $\mu$ V in V2 and QRS deflection positive in V3R or V1	PRINT REASON “RVH voltage criteria: R(V2) > 2.5mV, 1mo-15yr”
Age $\geq$ 1 month and maximum R amplitude/S amplitude < 1.2 in V6 and QRS deflection positive in V3R or V1	PRINT REASON “RVH voltage criteria: R/S(V6) < 1.2, 1mo-15yr” <b>NOTE:</b> Final comment is “Borderline ECG”
Maximum R amplitude/S amplitude < minimum V6 R/S amplitude ratio for Age <b>NOTE:</b> for age $\geq$ 3yr: and QRS deflection positive in V3R or V1	PRINT REASON “RVH voltage criteria: R/S(V6) < [Minimum V6 R/S amplitude ratio for age]”
Age < 5 days and maximum R amplitude > 2200 $\mu$ V in V3R or V1	PRINT REASON “RVH voltage criteria: R(V3R/V1) < 2.2mV, < 5day” <b>NOTE:</b> Final comment is “Borderline ECG”
Age $\geq$ 5 days and age < 30 days and maximum R amplitude > 2200 $\mu$ V in V3R or V1	PRINT REASON “RVH voltage criteria: R(V3R/V1) < 2.2mV, 5-30day”
Age $\geq$ 30 days and age < 16 years and maximum R amplitude > 1700 $\mu$ V in V3R or V1	PRINT REASON “RVH voltage criteria: R(V3R/V1) < 1.7mV, 1mo-15yr”
Maximum R amplitude/maximum S amplitude > maximum V3R/V1 R/S amplitude ratio for Age and R-amplitude in V3R/V1 $\geq$ 300 $\mu$ V	PRINT REASON “RVH voltage criteria: R/S(V3R/V1) > [Maximum V3R/V1 R/S amplitude ratio for Age]”
Age < 3 months and R' amplitude $\geq$ 2000 $\mu$ V and R' duration $\geq$ 20 ms and no S' in V3R or V1	PRINT REASON “RVH voltage criteria: R'(V3R/V1) > 2 mV, <3mo”

## Right Ventricular Hypertrophy Criteria (Continued)

IF	THEN
Age $\geq$ 2 months and Age < 1 year and R' amplitude $\geq$ 1600 $\mu$ V and R' duration $\geq$ 12 ms and no S' in V3R or V1	PRINT REASON "RVH voltage criteria: <i>R'(V3R/V1) &gt; 1.6 mV, 2-11mo</i> "
Age $\geq$ 1 year and R' amplitude > 1000 $\mu$ V and R' duration $\geq$ 12 ms and R' amplitude > R amplitude and R' amplitude > R amplitude and no S' in V3R or V1	PRINT REASON "RVH voltage criteria: <i>R'(V3R/V1) &gt; 1 mV, 1-15yr</i> "
Age < 5 days and QRS with only R-wave and R amplitude $\geq$ 1000 $\mu$ V in V3R or V1	PRINT REASON "RVH voltage criteria: <i>Pure R(V3R/V1) &gt; 1 mV, &lt;5day</i> "  <b>NOTE:</b> Final comment is "Borderline ECG"
Age $\geq$ 5 days and age < 30 days and QRS with only R-wave and R amplitude $\geq$ 1000 $\mu$ V in V3R or V1	PRINT REASON "RVH voltage criteria: <i>Pure R (V3R/V1) &gt; 1 mV, 5-30day</i> "
Age $\geq$ 30 days and QRS with only R-wave and R amplitude $\geq$ 500 $\mu$ V in V3R or V1	PRINT REASON "RVH voltage criteria: <i>Pure R (V3R/V1) &gt; 0.5 mV, 1mo-15yr</i> "
Age < 30 days and Q amplitude $\geq$ 70 $\mu$ V and Q duration $\geq$ 20 ms and R amplitude $\geq$ 500 and R amplitude > S amplitude in V3R or V1	PRINT REASON "RVH voltage criteria: <i>QR in V3R/V1, &lt; 1mo</i> "
Age $\geq$ 30 days Q amplitude $\geq$ 70 $\mu$ V and Q duration $\geq$ 20 ms and R amplitude $\geq$ 500 and R amplitude > S amplitude in V3R or V1	PRINT REASON "RVH voltage criteria: <i>QR in V3R/V1, 1mo-15yr</i> "



## Right Ventricular Hypertrophy Criteria (Continued)

IF	THEN
Age > 5 days and age < 5 years and T amplitude $\geq$ 100 $\mu$ V and T amplitude > 2 x STM amplitude in V3R or V1	PRINT REASON "RVH T wave criteria: <i>T upright in V3R/V1, 5day-4yr</i> "
Age $\geq$ 5 years and age < 9 years and T amplitude $\geq$ 150 $\mu$ V and T amplitude > 2 x STM amplitude in V3R or V1	PRINT REASON "RVH T wave criteria: <i>T upright in V3R/V1, 5-8yr</i> "  <b>NOTE:</b> <i>Final comment is "Borderline ECG"</i>

Please see pediatric criteria table for **V6 R/S Amplitude Ratio for Age** in Pediatric Reference Summary. Additionally, the following definitions are utilized: **STJ** = ST segment amplitude at QRS offset; **STM** = ST segment amplitude at ST segment midpoint; **STE** = ST segment amplitude at ST segment endpoint.

## Summary Statement

Depending upon which RVH criteria are satisfied, a summary statement reflecting the different criteria and their degree will be generated. Summary statements for RVH include the following:

Consider Right Ventricular Hypertrophy [Voltage Criteria Only]

Consider Right Ventricular Hypertrophy [T wave Changes]

Consider Right Ventricular Hypertrophy [Axis Criteria Only]

**NOTE:** *Final comment is "Borderline ECG"*

Probable Right Ventricular Hypertrophy [Voltage Criteria Only]

Probable Right Ventricular Hypertrophy [T wave Changes]

Right Ventricular Hypertrophy [Severe Voltage Criteria]

**NOTE:** *R/S(V3R/V1) > [Maximum V3R/V1 R/S amplitude ratio for Age], 1-11 mo; Pure R (V3R/V1) > 0.5 mV, 1mo-15yr; QR in V3R/V1, 1mo-15yr*

Right Ventricular Hypertrophy [T wave Changes & RAD for Age]

Right Ventricular Hypertrophy [Voltage & T wave Changes]

Right Ventricular Hypertrophy [Voltage & RAD for Age]

Right Ventricular Hypertrophy [Voltage & RAE]

Right Ventricular Hypertrophy [Voltage, RAD for Age & T wave Changes]

Consider Associated Right Ventricular Hypertrophy [R(V1) > 1.5 mV & LVH]

Consider Biventricular Hypertrophy [R+S > 6mV in 2 of V2-V4]

**NOTE:** *Final comment is "Borderline ECG"*

## Left Ventricular Hypertrophy

SKIP TEST IF
Test for Left Bundle Branch Block is positive, or Test for Right Bundle Branch Block is positive, or Test for Ventricular Conduction Delay is positive

### Criteria

Criteria statements for Left Ventricular Hypertrophy are printed only if the “Print reason” option on the electrocardiograph is turned on; otherwise, only the summary statements are printed.

IF	THEN
Q amplitude $\geq 600 \mu\text{V}$ and R amplitude $\geq 1000 \mu\text{V}$ in V5 or V6	PRINT REASON “LVH voltage criteria: $Q > 0.6\text{mV} \ \& \ R > 1 \text{ mV in V5/V6}$ ”  <b>NOTE:</b> Final comment is “Borderline ECG”
R amplitude $\geq 3000 \mu\text{V}$ in I, II, aVL, or aVF	PRINT REASON “LVH voltage criteria: $R > 3 \text{ mV in 1 of I/II/aVL/aVF}$ ”  <b>NOTE:</b> Final comment is “Borderline ECG”
S amplitude $\geq 3500 \mu\text{V}$ in V2	PRINT REASON “LVH voltage criteria: $S(V2) > 3.5 \text{ mV}$ ”  <b>NOTE:</b> Final comment is “Borderline ECG”
R amplitude $\geq 2300 \mu\text{V}$ in V6 and T amplitude $\leq 1/10$ of R amplitude in V6	PRINT REASON “LVH voltage criteria: $R(V6) > 2.3 \text{ mV} \ \& \ \text{small T}$ ”  <b>NOTE:</b> Summary statement is “Borderline ECG”
R amplitude $\geq 3000 \mu\text{V}$ in V6	PRINT REASON “LVH voltage criteria: $R(V6) \geq 3.0 \text{ mV}$ ”
R amplitude $\geq 2300 \mu\text{V}$ and Q amplitude $\geq 600 \mu\text{V}$ in V6	PRINT REASON “LVH voltage criteria: $R(V6) > 2.3 \text{ mV} \ \& \ Q(V6) > 0.6\text{mV}$ ”
R amplitude of V5 + S amplitude of V1 $\geq 3500 \mu\text{V}$ and T amplitude $\leq 1/10$ of R amplitude in V6	PRINT REASON “LVH voltage criteria: $S(V1) + R(V5) \geq 3.5 \text{ mV} \ \& \ \text{small T}$ ”  <b>NOTE:</b> Final comment is “Borderline ECG”
R amplitude of V5 + S amplitude of V1 $\geq 4500 \mu\text{V}$	PRINT REASON “LVH voltage criteria: $S(V1) + R(V5) \geq 4.5 \text{ mV}$ ”
STM $\leq -10 \mu\text{V}$ and down sloping ST-segment and T amplitude $\leq -50 \mu\text{V}$ in 2 of I/aVL/V4/V5/V6	PRINT REASON “LVH ST-T criteria: $ST < -0.01 \text{ mV} \ \& \ T < -0.05 \text{ mV in 2 of I/aVL/V4-6}$ ”

## Summary Statement

Depending upon which LVH criteria are satisfied, a summary statement reflecting the different criteria and their degree will be generated. Summary statements for LVH include the following:

Consider Left Ventricular Hypertrophy [Voltage Criteria Only]

Consider Left Ventricular Hypertrophy [T wave Changes]

*NOTE: Final comment is "Borderline ECG"*

Probable Left Ventricular Hypertrophy [Severe Voltage Criteria]

Probable Left Ventricular Hypertrophy [LAD for Age & ST-T Changes]

Left Ventricular Hypertrophy [Voltage Criteria & LAD for Age]

Left Ventricular Hypertrophy, Probably Severe, or Systolic Overload [Voltage Criteria & ST-T Rightward]

Consider Associated Left Ventricular Hypertrophy [ $Q > 0.1\text{mV}$  &  $R > 1\text{mV}$  in V6 &  $R+S > 3.5\text{ mV}$  in V4 with RVH]

Consider Biventricular Hypertrophy [ $R+S > 6\text{mV}$  in 2 of V2-V4]

*NOTE: Final comment is "Borderline ECG"*



# PEDIATRIC ST SEGMENT ABNORMALITIES

## ST Segment Elevation

SKIP TEST IF
The test for either right bundle branch block, left bundle branch block, is positive

### Criteria

IF	THEN
Lesser of STJ or STM $\geq 150 \mu\text{V}$ in 2 leads of V2,V3,V4,V5,V6	PRINT "Probable normal anterior ST variation" <b>REASON:</b> <i>ST &gt; 0.15 mV in 2 of V2-V5</i>
STJ/STM/STE all $\geq 150 \mu\text{V}$ in 1 of II/III/aVF and STJ/STM/STE all $\geq 100 \mu\text{V}$ in 2 of II/III/aVF	PRINT "Non-specific inferior ST segment changes, probably normal variation" <b>REASON:</b> <i>ST &gt; 0.15 mV in 1 of II/III/aVF</i>
Lesser of STJ or STM $\geq 150 \mu\text{V}$ in 2 leads of I/aVL/V6	PRINT "Probable normal anterolateral ST variation" <b>REASON:</b> <i>ST &gt; 0.15 mV in 2 of I/aVL/V6</i>
Lesser of STJ or STM $\geq 150 \mu\text{V}$ in 2 leads of I/aVL/V6 and Test is positive for any LVH criteria	PRINT "Anterolateral ST segment changes, probably secondary to LVH" <b>REASON:</b> <i>ST &gt; 0.15 mV in 2 of I/aVL/V6 &amp; LVH</i>

## ST Segment Depression

SKIP TEST IF
The test for either right bundle branch block, left bundle branch block, is positive

### Criteria

IF	THEN
STJ/STM $< -200 \mu\text{V}$ in 1 of V2/V3/V4/V5	PRINT "Non-specific anterior ST segment changes, consider subendocardial injury" <b>REASON:</b> <i>ST &lt; -0.2mV in 1 of V2-V5</i>
STJ/STM $< -200 \mu\text{V}$ in 1 of II/III/aVF	PRINT "Non-specific inferior ST segment changes, consider subendocardial injury" <b>REASON:</b> <i>ST &lt; -0.2mV in 1 of II/III/aVF</i>
STJ/STM $< -200 \mu\text{V}$ in 1 of I/aVL/V6 and STJ/STM $< -200 \mu\text{V}$ in 1 of V2/V3/V4/V5	PRINT "Non-specific anterolateral ST segment changes, consider subendocardial injury" <b>REASON:</b> <i>ST &lt; -0.2mV in 1 of I/aVL/V2-6</i>
STJ/STM $< -200 \mu\text{V}$ in 1 of I/aVL/V6 and STJ/STM $< -200 \mu\text{V}$ in 1 of V2/V3/V4/V5 a test for LVH is positive	PRINT "Anterolateral ST segment changes, probably secondary to LVH" <b>REASON:</b> <i>ST &lt; -0.2mV in 1 of I/aVL/V2-V6 and LVH</i>



# PEDIATRIC T WAVE ABNORMALITIES

## T Wave Abnormality, Ischemia

SKIP TEST IF
Age < 12

### Criteria

IF	THEN
T amplitude $\leq -10 \mu\text{V}$ or T' amplitude $\leq -10 \mu\text{V}$ in 2 of V1/V2/V3	PRINT "Anterior, non-specific T wave changes" <b>REASON:</b> $T < -0.01 \text{ mV}$ in 2 of V1-V3
T amplitude $\leq -10 \mu\text{V}$ or T' amplitude $\leq -10 \mu\text{V}$ in 2 of V1/V2/V3  and  T amplitude $\leq -10 \mu\text{V}$ or T' amplitude $\leq -10 \mu\text{V}$ in 1 of V4/V5	PRINT "Anterior, non-specific T wave changes" <b>REASON:</b> $T < -0.01 \text{ mV}$ in V1-V5
T amplitude $\leq -100 \mu\text{V}$ or T' amplitude $\leq -100 \mu\text{V}$ in 2 of V1/V2/V3	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -0.1 \text{ mV}$ in 2 of V1-V3
T amplitude $\leq -100 \mu\text{V}$ or T' amplitude $\leq -100 \mu\text{V}$ in 2 of V1/V2/V3  and  T amplitude $\leq -100 \mu\text{V}$ or T' amplitude $\leq -100 \mu\text{V}$ in 1 of V4/V5	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -0.1 \text{ mV}$ in V1-V5
T amplitude $\leq -500 \mu\text{V}$ or T' amplitude $\leq -500 \mu\text{V}$ in 2 of V1/V2/V3	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -0.5 \text{ mV}$ in 2 of V1-V3
T amplitude $\leq -500 \mu\text{V}$ or T' amplitude $\leq -500 \mu\text{V}$ in 2 of V1/V2/V3  and  T amplitude $\leq -500 \mu\text{V}$ or T' amplitude $\leq -500 \mu\text{V}$ in 1 of V4/V5	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -0.5 \text{ mV}$ in V1-V5
T amplitude $\leq -1000 \mu\text{V}$ or T' amplitude $\leq -1000 \mu\text{V}$ in 2 of V1/V2/V3	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -1 \text{ mV}$ in 2 of V1-V3

## T Wave Abnormality Ischemia Criteria (Continued)

IF	THEN
T amplitude $\leq$ -1000 $\mu$ V or T' amplitude $\leq$ -1000 $\mu$ V in 2 of V1/V2/V3  and  T amplitude $\leq$ -1000 $\mu$ V or T' amplitude $\leq$ -1000 $\mu$ V in 1 of V4/V5	PRINT "Anterior T wave changes" <b>REASON:</b> $T < -1$ mV in V1-V5
T amplitude $\leq$ -1000 $\mu$ V or T' amplitude $\leq$ -1000 $\mu$ V in 2 of V1/V2/V3  and  test is positive for any RVH criteria	PRINT "Consider anterior ischemia, probably secondary to RVH" <b>REASON:</b> $T < -1$ mV in V1-V3 & RVH
T amplitude $\leq$ -100 $\mu$ V or T' amplitude $\leq$ -100 $\mu$ V in 2 of II/III/aVF	PRINT "Non-specific inferior T wave changes" <b>REASON:</b> $T < -0.1$ mV in 2 of II/III/aVF
T amplitude $\leq$ -500 $\mu$ V or T' amplitude $\leq$ -500 $\mu$ V in 1 of II/III/aVF	PRINT "Non-specific inferior T wave changes" <b>REASON:</b> $T < -0.5$ mV in 1 of II/III/aVF
T amplitude $\leq$ -1000 $\mu$ V or T' amplitude $\leq$ -1000 $\mu$ V in 1 of II/III/aVF  or  T amplitude $\leq$ -500 $\mu$ V or T' amplitude $\leq$ -500 $\mu$ V in 2 of II/III/aVF	PRINT "Consider inferior ischemia" <b>REASON:</b> $T < -1$ mV in 1 of II/III/aVF or $T < -0.5$ mV in 2 of II/III/aVF
T amplitude $\geq$ 1000 $\mu$ V or T' amplitude $\geq$ 1000 $\mu$ V in 2 of I/aVL/V2/V3/V4/V5/V6  and  test for LVH is negative	PRINT "Non-specific anterolateral T wave changes, probably normal variant" <b>REASON:</b> $T > 1$ mV in 2 of I/aVL/V2-V6



## T Wave Abnormality Ischemia Criteria (Continued)

IF	THEN
<p>T amplitude <math>\leq -10 \mu\text{V}</math> or  T' amplitude <math>\leq -10 \mu\text{V}</math> in 1 of I/aVL/V6</p> <p>and</p> <p>T amplitude <math>\leq -10 \mu\text{V}</math> or  T' amplitude <math>\leq -10 \mu\text{V}</math> in 2 of V1/V2/V3</p> <p>and</p> <p>T amplitude <math>\leq -10 \mu\text{V}</math> or  T' amplitude <math>\leq -10 \mu\text{V}</math> in 1 of V4/V5</p> <p>and not</p> <p>T amplitude <math>\leq -1000 \mu\text{V}</math> or  T' amplitude <math>\leq -1000 \mu\text{V}</math> in 2 of V1/V2/V3</p>	<p>PRINT "Non-specific anterolateral T wave changes"</p> <p><b>REASON:</b> <math>T &lt; -0.01 \text{ mV}</math> in I/aVL/V2-V6</p>
<p>T amplitude <math>\leq -100 \mu\text{V}</math> or  T' amplitude <math>\leq -100 \mu\text{V}</math> in 1 of I/aVL/V6</p> <p>and</p> <p>T amplitude <math>\leq -100 \mu\text{V}</math> or  T' amplitude <math>\leq -100 \mu\text{V}</math> in 2 of V1/V2/V3</p> <p>and</p> <p>T amplitude <math>\leq -100 \mu\text{V}</math> or  T' amplitude <math>\leq -100 \mu\text{V}</math> in 1 of V4/V5</p> <p>and not</p> <p>T amplitude <math>\leq -1000 \mu\text{V}</math> or  T' amplitude <math>\leq -1000 \mu\text{V}</math> in 2 of V1/V2/V3</p>	<p>PRINT "Non-specific anterolateral T wave changes"</p> <p><b>REASON:</b> <math>T &lt; -0.1 \text{ mV}</math> in I/aVL/V2-V6</p>

## T Wave Abnormality Ischemia Criteria (Continued)

IF	THEN
<p>T amplitude <math>\leq</math> -100 <math>\mu</math>V or  T' amplitude <math>\leq</math> -100 <math>\mu</math>V in 1 of I/aVL/V6  and  T amplitude <math>\leq</math> -100 <math>\mu</math>V or  T' amplitude <math>\leq</math> -100 <math>\mu</math>V in 2 of V1/V2/V3  and  T amplitude <math>\leq</math> -100 <math>\mu</math>V or  T' amplitude <math>\leq</math> -100 <math>\mu</math>V in 1 of V4/V5  and not  T amplitude <math>\leq</math> -1000 <math>\mu</math>V or  T' amplitude <math>\leq</math> -1000 <math>\mu</math>V in 2 of V1/V2/V3  and  a test is positive for any LVH criteria</p>	<p>PRINT "Consider anterolateral ischemia, probably secondary to LVH"  <b>REASON:</b> <math>T &lt; -0.1</math> mV in I/aVL/V2-V6 &amp; LVH</p>
<p>T amplitude <math>\leq</math> -500 <math>\mu</math>V or  T' amplitude <math>\leq</math> -500 <math>\mu</math>V in 1 of I/aVL/V6  and  T amplitude <math>\leq</math> -500 <math>\mu</math>V or  T' amplitude <math>\leq</math> -500 <math>\mu</math>V in 2 of V1/V2/V3</p>	<p>PRINT "Non-specific anterolateral T wave changes, probable ischemia"  <b>REASON:</b> <math>T &lt; -0.5</math> mV in 1 of I/aVL/V2-V6</p>
<p>T amplitude <math>\leq</math> -500 <math>\mu</math>V or  T' amplitude <math>\leq</math> -500 <math>\mu</math>V in 1 of I/aVL/V6  and  T amplitude <math>\leq</math> -1000 <math>\mu</math>V or  T' amplitude <math>\leq</math> -1000 <math>\mu</math>V in 2 of V1/V2/V3</p>	<p>PRINT "Consider anterolateral ischemia"  <b>REASON:</b> <math>T &lt; -1</math> mV in 1 of I/aVL/V2-V6</p>

# PEDIATRIC TRICUSPID ATRESIA

---

## Tricuspid Atresia

### Criteria

IF	THEN
Left axis deviation and Left ventricular hypertrophy and Right atrial enlargement	PRINT "Consider tricuspid atresia" <b>REASON:</b> <i>RAE + LAD + LVH</i>



# PEDIATRIC ENDOCARDIAL CUSHION DEFECT

---

## Endocardial Cushion Defect

### Criteria

IF	THEN
QRS Axis > -170 or QRS Axis < -30 and Q amplitude $\geq 80 \mu\text{V}$ and R amplitude $\geq 100 \mu\text{V}$ in aVL and Test for RVH or RBBB is true	PRINT "Consider endocardial cushion defect" <b>REASON:</b> QRS -30 to -170, RVH or RSR' in V1



# PEDIATRIC ATRIAL SEPTAL DEFECT

---

## Atrial Septal Defect

### Criteria

IF	THEN
QRS Axis $> 0$ and QRS Axis $\leq 180$ and RSR' pattern in V1	PRINT "Consider atrial septal defect" <b>REASON:</b> <i>QRS 1-180, RSR' in V1</i>





# PEDIATRIC QT INTERVAL ABNORMALITIES

---

## QT Prolongation

### Criteria

IF	THEN
QTc > 450	PRINT "Long QT Interval" <b>REASON:</b> <i>QTc &gt; 450</i>

## QT Shortening

### Criteria

IF	THEN
QTc < 340 and Ventricular rate < 140	PRINT "Short QT Interval" <b>REASON:</b> <i>QTc &lt; 340</i>



# PEDIATRIC BRUGADA

---

## Brugada

SKIP TEST IF
QRS duration $\geq$ 110 ms

### Criteria

IF	THEN
STJ $\geq$ 100 $\mu$ V and the lesser of STM or STE $<$ STJ and T amplitude $>$ the lesser of STM or STE in some of V1/V2/V3 and T amplitude $<$ 0 in some of V1/V2/V3	PRINT "Type 3 Brugada pattern (non-diagnostic)"
STJ $\geq$ 200 $\mu$ V and the lesser of STM or STE $<$ STJ and the lesser of STM or STE $>$ 100 $\mu$ V and T amplitude $>$ the lesser of STM or STE in some of V1/V2/V3	PRINT "Type 2 Brugada pattern (non-diagnostic)"
STJ $\geq$ 200 $\mu$ V and the lesser of STM or STE $<$ STJ and T amplitude $<$ 0 in some of V1/V2/V3	PRINT "Type 1 Brugada pattern (non-diagnostic)"



## REFERENCE SUMMARY

---

### Age Tables

The following tables should be used in reference to parameters that are stated as minimum or maximum “for age”. In the following, **d** indicates days, **mo** indicates months, and **yr** indicates years.

### QRS Axis for Age

Age	QRS Axis Minimum (Left Axis Deviation Criteria)	QRS Axis Maximum (Right Axis Deviation Criteria)
< 6d	60	180
6 – 30d	60	160
1 – 2mo	40	135
3 – 5mo	20	135
6mo – 15yr	0	135

### QRS Duration for Age

Age	QRS Duration
< 1yr	99ms
1 – 15yr	109ms

### Prolonged PR Duration, Bradycardia, and Tachycardia for Age

Age	PR Duration (ms)	Bradycardia (bpm)	Tachycardia (bpm)
< 1d	129	94	145
< 8d	129	100	175
< 1mo	129	115	190
< 3mo	139	124	190
< 1yr	139	110	178
< 3yr	159	98	163
< 5yr	159	65	132
< 8yr	169	65	115
< 12yr	179	60	107
< 16yr	179	60	102
≥ 16yr	NEED	55	99

### V6 R/S Amplitude Ratio for Age

Age	V6 R/S Amplitude Ratio
1 – 3mo	0.5
4 – 11mo	0.7
1 – 2yr	0.8
3 – 15yr	0.9

### V1/V3R R/S Amplitude Ratio for Age

Age	V1/V3R R/S Amplitude Ratio
1 – 3mo	7
4 – 11mo	4.5
1 – 2yr	3
3 – 7yr	2.3
8 – 15yr	2



# VERITAS RESTING ECG INTERPRETATION EVALUATION

---

## METHOD

### Introduction and General Methodology

To test the analysis program 558 12-lead ECGs were randomly collected from adult patients in a clinical and hospital setting over a two (2) month period. These ECGs were collected and then stored in digital form. Separately, 553 pediatric 15 lead ECG's (standard 12 leads plus V3R, V4R and V7) were collected from various pediatric hospital settings.

In addition, consecutive 568 ECG's were added, half from a hospital digital ECG archive, half from an ambulance service, in order to increase the statistical reliability for the major rhythm categories sinus rhythm, atrial fibrillation and atrial flutter.

To test the criteria, ECGs from the data bases were submitted to a doctor to interpret as one would when reading a standard ECG available in a typical heart station. In addition, the same ECGs were interpreted by the Mortara Instrument Veritas Analysis program running on a personal computer.

No reinterpretation was allowed by the doctor or the electrocardiograph.

Some categories of statements have been tested separately with different databases and methodologies, specifically pediatric ventricular hypertrophy statements and electronic pacemaker statements. The reasons and methods are explained below.

### Pediatric Ventricular Hypertrophy

Left and Right Ventricular Hypertrophy (LVH and RVH) are the most common ECG interpretations in a typical pediatric cardiology population. Criteria for hypertrophy are complex, sometimes controversial and highly age dependent. This is why the performance of the program for Left and Right Hypertrophy has been tested differently and more extensively. Approximately 1300 15-lead ECG's (standard 12 leads plus V3R, V4R and V7) were randomly collected in various pediatric cardiology centers.

To test the criteria, ECGs from this database were submitted to a cardiologist without automatic interpretation (blind reading) and in a standard 3x5 format at 10 mm/mV and 25 mm/s. The cardiologist was asked to divide the ECG's in 3 groups: "No RVH", "Possible RVH" and "RVH". Subsequently, the same cardiologist was represented with the ECGs but now had to divide the ECGs in "No LVH", "Possible LVH" and "LVH". In addition, the same ECG's were interpreted by the VERITAS Pediatric ECG Interpretation algorithm.

ECGs with a wide QRS (Right Bundle Branch Block, Left Bundle Branch Block, Ventricular Conduction Delay and Ventricular Pre-Excitation) were excluded from the analysis. The VERITAS Algorithm omits the RVH and LVH calls in these cases because criteria for hypertrophy in the presence of abnormal intra-ventricular conduction are poorly defined.

ECGs with an erroneous order of the V-leads (for instance V7 at the place of V3R) were also excluded, leading to 1,174 included ECGs in total. Subsequently, the VERITAS algorithm was run again using only the standard 12 leads.

The tables below indicate the performance of the VERITAS program, using as "truth" both the possible and definite hypertrophy groups from the cardiologist, in confrontation with any hypertrophy call of the program. Note that the "truth" was always defined on the full 15 lead ECG.

## Pacemaker Detection

The acquisition method of the aforementioned databases did not allow for adequate testing of the detection of artificial pacemaker rhythms because of the low prevalence of some pacemaker types and because of insufficient quality of the pacemaker pulse registration in the older data. Instead, 69 ECG's from patients with various types of pacemaker stimulation (6 atrial only, 48 ventricular, 15 atrial and ventricular; about 25% also showed intrinsic rhythm) were collected from a pacemaker evaluation center. These ECG's were used to establish the sensitivity of the VERITAS program (more precisely, the percentage of undetected pacemaker rhythms).

A large database with circa 7000 ECG's from various institutions, was used to measure the number of false positive pacemaker detections: all ECG's with an "Artificial Pacemaker" statement from the VERITAS program were reviewed by an expert. In this way it was possible to establish the percentage of false positives. The statistical measurements (see below) were subsequently calculated on the basis of a population with 1% pacemaker ECG's.

## Comparison by Categories

For purposes of determining specificity, sensitivity, positive and negative predictive accuracy, statements have been grouped into categories. This has been done for various reasons: A higher number per category increases statistical significance; severity and probability statements (minimal, marked, possible, probable) are not well defined and highly subjective; some electrocardiographic regions (septal, anteroseptal, anterior, anterolateral and lateral) overlap and are not well defined; tachycardia, bradycardia and "normal" rate differ only in heart rate, while the algorithm that establish the statements is the same; the VERITAS sometimes issues a generic statement (e.g. supraventricular, uncertain) when an abnormality is detected, while the cardiologist will usually attempt to be more specific.

Some statements exist only for adult or pediatric populations and have been tested only in those. Some statements have very different meaning or prevalence in pediatric or adult population, and have been tested separately.

Below is the list of categories that have been used for statistical analysis, and the VERITAS statements that are grouped into them:

- Sinus Rhythm
  - Normal Sinus Rhythm
  - Sinus Bradycardia
  - Sinus Tachycardia
- Atrial Fibrillation
- Atrial Flutter
- Miscellaneous Rhythms
  - Ectopic Atrial Rhythm
  - Ectopic Atrial Tachycardia
  - Ectopic Atrial Bradycardia
  - Junctional Rhythm
  - Junctional Tachycardia
  - Junctional Bradycardia
  - Idioventricular Rhythm
  - Ventricular Tachycardia
  - Supraventricular Rhythm
  - Supraventricular Tachycardia
  - Supraventricular Bradycardia
  - Uncertain Irregular Rhythm
  - Uncertain Regular Rhythm



- Supraventricular Premature Complexes
  - With Occasional Atrial Premature Complexes
  - With Frequent Atrial Premature Complexes
  - With Occasional Supraventricular Premature Complexes
  - With Frequent Supraventricular Premature Complexes
- Ventricular Premature Complexes
  - With Occasional Ventricular Premature Complexes
  - With Frequent Ventricular Premature Complexes
  - With Occasional Ectopic Premature Complexes
  - With Frequent Ectopic Premature Complexes
- Atrial Electronic Pacemaker
- Ventricular Electronic Pacemaker
- High degree AV-block
  - With second degree AV-block type Mobitz 1 (Wenckebach)
  - With second degree AV-block type Mobitz 2
  - With high degree AV-block
- Prolonged PR-Interval (First Degree AV-block)
- Short PR-interval (adult only)
- Right Atrial Enlargement
  - Possible Right Atrial Enlargement
  - Right Atrial Enlargement
- Left Atrial Enlargement
  - Possible Left Atrial Enlargement
  - Left Atrial Enlargement
- Right Axis Deviation
  - Borderline Right Axis Deviation
  - Marked Right Axis Deviation
- Left Axis Deviation
  - Borderline Left Axis Deviation
  - Marked Left Axis Deviation
- Low QRS Voltage (adult only)
  - Low QRS Voltage In Extremity Leads
  - Low QRS Voltage In Precordial Leads
  - Low QRS Voltage
  - S1-S2-S3 Pattern, Consistent With Pulmonary Disease, Rvh, or Normal Variant
  - Pattern Consistent With Pulmonary Disease
- Right Bundle Conduction
  - Right Bundle Branch Block
  - Right Bundle Branch Block, plus possible right ventricular hypertrophy
  - Note: moderate right conduction delays have not been considered

- Nonspecific intraventricular conduction block  
     Note: moderate conduction delays have not been considered
- Left bundle branch block  
     Note: moderate left conduction delays and fascicular blocks have not been considered
- Right Ventricular Hypertrophy  
     Possible Right Ventricular Hypertrophy  
     Probable Right Ventricular Hypertrophy  
     Right Ventricular Hypertrophy  
     Right Ventricular Hypertrophy and ST-T Change
- Left Ventricular Hypertrophy  
     Minimal Voltage Criteria for LVH, Consider Normal Variant  
     Moderate Voltage Criteria for LVH, Consider Normal Variant  
     Voltage Criteria for LVH  
     Possible Left Ventricular Hypertrophy  
     Probable Left Ventricular Hypertrophy  
     Left Ventricular Hypertrophy and S-T Change

***NOTE:** separate tables are compiled for Ventricular Hypertrophy for Adults, Pediatric 12-lead ECG's and Pediatric 15-lead ECG's*

- Inferior Infarction (adult only)  
     All inferior infarction statements
- Anterior Infarction (adult only)  
     All septal, anterior, lateral, anteroseptal and anterolateral infarction statements
- ST-T changes - adult  
     All adult ST-depression and T-wave abnormality statements
- ST-T changes - pediatric  
     All pediatric ST-depression and T-wave abnormality statements
- Prolonged QT
- Consider Endocardial Cushion Effect (pediatric only)

## Results

Results are presented in two different forms. In order to more clearly view the positive and negative calls by the physician and the Mortara Instrument VERITAS Analysis Program, the following tables present data in a 2 x 2 truth matrix format (Table 1 and 2). Below, summary statistical measurements like sensitivity and specificity are given (Table 3 and 4). For this presentation, the categories have been divided in two groups: rhythm statements and statements based on waveform morphology.

## Definitions

In the matrix format shown, the Physician Statement is used as the gold standard against which the Mortara Instrument VERITAS ECG Analysis Program is compared.

		Mortara Instrument VERITAS Analysis Program	
		+	-
Physician Statement	+	True Positive	False Negative
	-	False Positive	True Negative

Specific definitions for each of the terms used above are as follows:

**True Positive:**  
(TP)

A true positive is called when the analysis program agrees with the positive diagnostic statement made by the physician, i.e., true positive call by the analysis program.

**True Negative:**  
(TN)

A true negative is called when the analysis program agrees with the negative diagnostic statement made by the physician, i.e., the condition under question is not called by either the analysis program or the physician.

**False Positive:**  
(FP)

A false positive occurs when the analysis program appends the diagnostic statement to the ECG in question whereas the physician indicates that the condition did not exist, i.e., a false positive call by the analysis program.

**False Negative:**  
(FN)

A false negative occurs when the physician appends the diagnostic statement to the ECG in question whereas the analysis program indicates that the condition did not exist, i.e., a false negative call by the analysis program.

In summary, True Positive and True Negatives are correct diagnostic statements made by the analysis program since they truly reflect the positive and negative calls made by the physician. False Positives and False Negatives occur when the analysis program calls do not agree with the physician statement. A false positive, in effect, overcalls a particular diagnostic statement whereas a false negative undercalls. The prevalence of the condition in the databases used can be determined by summing the True Positive and False Negative numbers.

In addition, the values for sensitivity, specificity and predictive accuracy are presented in table form following the analysis matrices. True Positives, True Negatives, False Positives and False Negatives have been used to calculate the Sensitivity, Specificity and the Predictive Accuracy.

Formulas used for calculating the above values are:

$$\text{Sensitivity} = \frac{\text{TP}}{\text{TP} + \text{FN}} \qquad \text{Specificity} = \frac{\text{TN}}{\text{TN} + \text{FP}}$$

$$\text{Positive Predictive Accuracy} = \frac{\text{TP}}{\text{TP} + \text{FP}}$$

$$\text{Negative Predictive Accuracy} = \frac{\text{TN}}{\text{TN} + \text{FN}}$$

Table 1, Rhythm Criteria Truth Matrices

		Sinus Rhythm				Atrial Fibrillation				Atrial Flutter	
		+	-			+	-			+	-
+	+	1461	53	+	+	99	10	+	+	5	5
	-	15	150		-	-	16		1554	-	-
		Miscellaneous Rhythms				High Degree AV-Block				Ventricular Preexcitation	
		+	-			+	-			+	-
+	+	28	0	+	+	2	5	+	+	7	8
	-	26	1057		-	-	0		1104	-	-

Table 1, Rhythm Criteria Truth Matrices (Continued)

<b>Premature Ventricular Complexes</b>			<b>Supraventricular Premature Complexes</b>	
+	-		+	-
58	3	+	25	3
14	1036	-	14	1075

<b>Atrial Electronic Pacemaker</b>			<b>Ventricular Electronic Pacemaker</b>	
+	-		+	-
0.38 %	0.38 %	+	0.81 %	0.19 %
0.04 %	98.96 %	-	0.04 %	98.96 %

Table 2, Contour Criteria Truth Matrices

		Prolonged PR-Interval		Short PR-Interval (Adult Only)		Right Atrial Enlargement	
		+	-	+	-	+	-
+		61	11	12	6	25	6
	-	6	1033	4	536	13	1067
		Left Atrial Enlargement		Right Axis		Left Axis	
		+	-	+	-	+	-
+		55	13	30	4	83	12
	-	21	1022	6	1071	5	1011
		Low QRS Voltage (Adult Only)		Right Bundle Conduction		Nonspecific Conduction Abnormality	
		+	-	+	-	+	-
+		8	6	52	6	11	2
	-	3	541	6	1047	15	1083

Table 2, Contour Criteria Truth Matrices (Continued)

		Left Bundle Conduction				Right Ventricular Hypertrophy (Adult Only)				Left Ventricular Hypertrophy (Adult Only)	
		+	-			+	-			+	-
+		12	2	+		5	5	+		123	4
-		2	1095	-		1	547	-		10	421

		Right Ventricular Hypertrophy (Pediatric 12 Lead)				Left Ventricular Hypertrophy (Pediatric 12 Lead)				Right Ventricular Hypertrophy (Pediatric 15 Lead)	
		+	-			+	-			+	-
+		113	59	+		126	29	+		137	35
-		52	950	-		51	968	-		74	928

		Left Ventricular Hypertrophy (Pediatric 15 Lead)	
		+	-
+		127	28
-		51	968



Table 2, Contour Criteria Truth Matrices (Continued)

		Anterior Infarction (Adult Only)		Inferior Infarction (Adult Only)		ST-T Changes Adult	
		+	-	+	-	+	-
+		52	20	70	11	133	22
		4	482	14	463	60	896
-							

  

		ST-T Changes Pediatric		Prolonged QT		Endocardial Cushion Effect (Pediatric)	
		+	-	+	-	+	-
+		18	3	20	0	6	1
		3	528	14	1077	6	540
-							

Table 3, Sensitivity, Specificity and Predictive Accuracies, Rhythm Criteria

RHYTHM CRITERIA				
DIAGNOSTIC STATEMENT	SENSITIVITY	SPECIFICITY	POS PREDICTIVE ACCURACY	NEG PREDICTIVE ACCURACY
Sinus Rhythm	96.5	90.9	99.0	73.9
Atrial Fibrillation	90.8	99.0	86.1	99.4
Atrial Flutter	29.4	99.9	71.4	99.3
Miscellaneous Rhythms	100	97.6	51.9	100
High Degree AV-Block	28.6	100	100	99.5
Ventricular Preexcitation	46.7	100	100	99.3
Ventricular Premature Complexes	95.1	98.7	80.6	99.7
Supraventricular Premature Complexes	73.5	99.8	92.6	99.2
Atrial Electronic Pacemaker	38.0	100	90.5	99.4
Ventricular Electronic Pacemaker	81.0	100	95.3	99.8

Table 4, Sensitivity, Specificity and Predictive Accuracies, Contour Criteria

CONTOUR CRITERIA				
DIAGNOSTIC STATEMENT	SENSITIVITY	SPECIFICITY	POS PREDICTIVE ACCURACY	NEG PREDICTIVE ACCURACY
Prolonged PR-Interval	84.7	99.4	91.0	98.9
Short PR-interval (Adult)	66.7	99.3	75.0	98.9
Right Atrial Enlargement	80.6	98.8	100	99.3
Left Atrial Enlargement	80.9	98.0	72.4	98.7
Right Axis	88.2	99.4	83.3	99.6
Left Axis	87.4	99.5	94.3	98.8
Low QRS Voltage (Adult)	57.1	99.7	72.7	99.5
Right Bundle Conduction	89.7	99.4	89.7	99.4
Nonspecific Conduction Abnormality	84.6	98.6	42.3	99.8
Left Bundle Conduction	85.7	99.8	85.7	99.8
Right Ventricular Hypertrophy, Adult	50.0	99.8	83.3	99.1
Left Ventricular Hypertrophy, Adult	96.9	97.7	92.5	99.1
Right Ventricular Hypertrophy, Pediatric 12 Lead	65.7	94.8	68.5	94.2
Left Ventricular Hypertrophy, Pediatric 12 Lead	81.3	95.0	71.2	97.1
Right Ventricular Hypertrophy, Pediatric 15 Lead	79.7	92.6	64.9	96.4
Left Ventricular Hypertrophy, Pediatric 15 Lead	81.9	95.0	71.3	97.2
Inferior Infarction	86.4	98.6	83.3	98.9
Anterior Infarction	72.2	99.8	92.9	98.1
ST-T Changes, Adult	85.8	93.7	68.9	97.6
ST-T Changes, Pediatric	85.7	99.7	85.7	99.7
Prolonged QT	100	98.7	58.8	100
Endocardial Cushion Effect (Pediatric)	85.7	99.5	50.0	99.9

## Interval Measurements

The global PR-interval, QRS-duration and QT-interval are measured using the “median beat”, using all available leads. The first and last wave of the QRS for individual leads start and end with the global onset and offset of the QRS, therefore iso-electric segments before the Q-wave and after the S-wave may be included in the Q or S duration measurements of the program.

The interval measurements have been tested according to IEC 60601-2-51 (2003) on 100 ECG’s with established “truth” A positive difference means that the Veritas measurement is bigger than the truth.

Table 5, Accuracy of Interval Measurements

Global measurement	Acceptable mean difference	Measured mean difference	Dimensions in ms	
			Acceptable standard deviation	Measured standard deviation
PQ.interval	± 10	2.1	10	7.2
QRS-duration	± 10	-0.4	10	5.9
QT-interval	± 25	-7.8	30	10.6

The stability of the measurements in conditions of noise has been measured according to IEC 60601-2-51 (2003), by adding high frequency, line frequency and base-line artifact and comparing the results with the measurements on the same ECG’s without noise. Results were as follows:

Table 6, Stability of Interval Measurements Against Noise

Global measurement	Type of added noise	Disclosed differences	
		Mean ms	Standard deviation ms
P-duration	High frequency	1.50	3.21
P-duration	Line frequency	0.63	2.00
P-duration	Base-line	0.13	1.46
QRS-duration	High frequency	-0.38	1.51
QRS-duration	Line frequency	0.13	0.99
QRS-duration	Base-line	-0.25	1.28
QT-interval	High frequency	0.25	1.58
QT-interval	Line frequency	0.13	1.55
QT-interval	Base-line	-0.13	0.99